ISSN: 2093-5374 / E-ISSN: 2093-5382

A Novel Algorithm for Respiration Rate Detection Using Deep Learning and Real-Time Sensor Data

Dr.J. Sylvia Grace[1*], M. Ulagammai[2], Rajat Saini[3], Harsimrat Kandhari[4], R. Rajalakshmi[5], and Dr.M.K. Prathiba[6]

[1*]Department of Computer Science and Engineering, Sathyabama Institute of Science and Technology, Chennai, India. ssylviagrace@gmail.com, https://orcid.org/0000-0001-9044-9474

[2]Saveetha Engineering College, Chennai, India. ulagammaimeyyappan@gmail.com, https://orcid.org/0000-0002-4771-1593

[3]Centre of Research Impact and Outcome, Chitkara University, Rajpura, Punjab, India. rajat.saini.orp@chitkara.edu.in, https://orcid.org/0009-0009-7750-9896

[4]Chitkara Centre for Research and Development, Chitkara University, Himachal Pradesh, India. harsimrat.khandari.orp@chitkara.edu.in, https://orcid.org/0009-0005-8187-3077

[5]Department of ECE, Panimalar Engineering College, Chennai, India. rajeeramanathan@gmail.com, https://orcid.org/0000-0002-4399-4071

[6]Department of Electronics and Communication Engineering, ATME College of Engineering, Mysuru, India. dr.prathibamk_ec@atme.edu.in, https://orcid.org/0000-0002-8865-189X

Received: June 04, 2025; Revised: July 14, 2025; Accepted: September 01, 2025; Published: September 30, 2025

Abstract

One crucial parameter for monitoring a person's health is the respiratory rate (RR). RR detection has been performed both automatically and manually in the past. We have recently focused our efforts on solving the RR detection problem with SREC machines and ML algorithms, and it appears promising that we will soon have a solution for completely automatic RR detection. Nevertheless, some deficiencies exist in the previous methods, such as the requirement for specialized sensors, low accuracy, or incompleteness of the real-time function. Motivated by these drawbacks, in the following letter, we propose a deep-learning solution to detect abnormal waveform segments in real-time from in-situ sensor data. To utilize the sensor data, we train Convolutional Neural Networks (CNNs) to extract expressive features and reliably detect RR intervals. Our approach is low-cost, simple to deploy, and able to scale without the need for many custom sensors or additional hardware. It also includes a feature selection of the input data to decrease the size of the input data while enhancing the detection accuracy. We applied our algorithm to real-time sensor data from an experiment conducted on various individuals and compared its performance with that of previous approaches. The results show that our method not only outperforms related algorithms in terms of accuracy but also in terms of real-time capability.

Journal of Wireless Mobile Networks, Ubiquitous Computing, and Dependable Applications (JoWUA), volume: 16, number: 3 (September), pp. 334-353. DOI: *10.58346/JOWUA.2025.I3.019*

*Corresponding author: Department of Computer Science and Engineering, Sathyabama Institute of Science and Technology, Chennai, India.

Keywords: Respiratory Rate, Time-Consuming, Limited Accuracy, Convolutional Neural Networks, Sensor Data, Feature Selection.

1 Introduction

It's an automatic way to track how quickly somebody breathes. The fundamental principle of this algorithm is that it acquires signals from a sensor mounted on the chest or finger to measure the displacement of the chest or finger while breathing (Shuzan et al., 2021). It then processes these signals to estimate the breathing rate of the person. The algorithm's process comprises three stages: signal acquisition, processing, and respiration rate analysis. The sensor records chest or finger movements caused by respiration and translates them into electrical signals - this is the signal acquisition step (Kumar et al., 2022). These signals are then sent to a computer for further processing. The raw signals contain noise, and they are assumed to be noise-free for analysis, as they are amplified and filtered during the signal processing step to provide the real data (Stankoski et al., 2022). It is comprised of high-pass and low-pass filters using different digital signal processing methods. Finally, Fourier transforms can be employed to transform the above signal into the frequency domain, allowing for the recognition of components associated with the frequency of breathing (Selvakumar et al., 2022). Finally, the estimated frequency components are used to calculate the respiration rate. The algorithm utilizes one or more machine-learning methods to produce an accurate prediction of the respiratory rate based on the frequency components. One of the advantages of the algorithm is that it's customizable, even across different individuals or average breath patterns. It is capable of handling irregular and noisy signals without introducing errors. This is achieved by employing several feature selection and extraction techniques to identify relevant signal features on which to process the signal to estimate the respiration rate e.g., (Zhang et al., 2023). Moreover, it has also been proven capable of real-time monitoring of respiratory rates, rendering it applicable for both clinical and in-home monitoring. It could potentially be incorporated into wearable devices for continuous measurement of respiration without bulky equipment (Shu et al., 2022). An alternative to this method is a computer model that estimates an individual's rate of breathing based on biological signals, as described in a Novel Algorithm for Respiration Rate Detection. Although this makes the method skillful for these reasons, it did not prevent additional issues, as covered in other respects, particularly in terms of informativeness and reliability. This is achieved by applying several feature selection algorithms, taking the output of the pipeline, and identifying the significant signal features over which signal processing for respiration rate estimation can be used, i.e., (Islam et al., 2023). It has also been demonstrated that the algorithm is robust for real-time respiration rate monitoring, covering both clinical and home users. Its form factor may allow it to be embedded in wearable devices to monitor respiratory rates, even without bulky machine devices that continuously monitor. An alternative approach to the above is a computational model for estimating respiration frequency in a person using biomarkers, known as a novel algorithm for respiration rate detection. Since the approach is highly skilled for these reasons, several difficulties, however, arose regarding instructiveness and reliability. The main technical difficulty is that this is one of the advantages of this algorithm; even for different individuals or normal breaths, this algorithm can be applied. It does not produce errors in the presence of irregular and noisy signals. This is achieved by using several feature selection and extraction techniques to identify the significant signal features where breathing rate signal processing can be performed, as described in (Shinde et al., 2022). Moreover, the algorithm achieved good precision for respiration rate, making it suitable for both clinical and home use. Due to its design, it can be embedded in wearable sensing devices, allowing for the monitoring of respiration rates over long periods without the need for laboratory instrumentation. A different approach is a model-based method for estimating a person's respiration rate, which predicts the respiration rate in terms of

biological signals, as described in a novel algorithm for Respiration Rate Detection. Although this approach is highly skillful, it also contributed in other respects, namely in terms of the complexity of learning and terms of quadraticity and reliability (Alnaggar et al., 2023). The main contribution of the research has the following:

- Development of a novel and computationally efficient respiratory rate detection algorithm: In the present study, the authors introduce a method for real-time measurement of precise respiratory rates from non-invasive sensor signals obtained from photoplethysmography (PPG) and/or electrocardiography (ECG). In this work, we adopted the Compressive Sampling Theory (CST) based L1 minimization criterion to monitor the Lexed in the respiration Signal with a proposed algorithm that combined both statistical and wavelet domain techniques to exploit and filter certain characteristics of Lexed signal to combat some limitations exhibited by the classical process of the respiration detection such as sensitivity to noise.
- The effectiveness of the proposed algorithm is confirmed using various datasets, including private and clinical datasets. The method indicating the viability of the proposed RR detection method as an upgrade to existing RR detection tools.
- Real-time deployment. One of the key contributions of the present work is the real-time deploy ability of the R-detection algorithm. This is achieved using fast and computationally efficient signal processing methods (which can be implemented at a low cost with portable devices) that make the technology suitable, for example, for remote health monitoring and other healthcare applications. It contributes to the high practical feasibility and applicability of the proposed algorithm.
- The remaining part of the research has the following chapters. Chapter 2 describes the recent works related to the research. Chapter 3 describes the proposed model, and chapter 4 describes the comparative analysis. Finally, chapter 5 shows the result, and chapter 6 describes the conclusion and future scope of the research.

2 Related Words

He et, al., 2022 have discussed A real-time respiration monitoring and classification system based on a combination of depth cameras and radars to accurately track and analyze the respiration of a human being. This method involves acquiring body signals and can be used for helpful medical monitoring, sleep analysis, or even biofeedback training without the need for invasive devices. Alharbi, et al., 2021 have discussed TXT - Predicting heart rate in real-time using state-of-the-art deep learning and stream processing technologies. This enables the understanding of real-time measurement and analysis of physiological data for HR prediction and abnormality detection. This can be accomplished with sophisticated algorithms and rapid data processing capabilities, enabling doctors to learn much more about themselves and their patients. Bokka, et al., 2021 have discussed Artificial intelligence, Specifically Deep learning, is a type of artificial intelligence that has been used to classify real-time respiration signals recorded through MoSSe QD-based flexible sensors. This method processes the data obtained by the sensor to perform an efficient classification of different types of breathing, thereby offering fast and reliable respiratory monitoring for various applications. Filosa, et al., 2022 have discussed A meta-learning program, similar to a type of machine learning, learns how to learn and then applies that knowledge. In this context, it forecasts flow patterns when wearable sensors register Fiber Bragg Grating (FBG) signals. It has shown satisfactory performance even under unconstrained conditions, making it a potential tool for real-time breath monitoring. Wang et, al., 2021 have discussed The ELC model - a model for describing the health of infants during perioperative care. This involves

monitoring the heart rate and respiratory rate, as well as vital signs of the cardiovascular and respiratory systems. Monitoring of these vital signs is crucial for the safety and welfare of the newborn during surgery. Dai et al., 2021 have discussed Heart rate is a key metric for wellness. Light-based sensors, such as photoplethysmography (PPG), for measuring respiratory rate can now also be found in smartwatches. It offers non-contact, continuous, and stable monitoring of breathing rate, allowing for daily assessment of respiratory health. Bandyopadhyay et, al., 2024 have discussed the early-stage detection of silent hypoxia in COVID-19 pneumonia is possible using deep learning and the Internet of Things (IoT). Through wearables and medical-grade sensors, algorithms can reliably detect low oxygen levels in patients, sending alerts to healthcare providers so that interventions can be made, potentially saving a life or even preventing severe illness or death. Chowdhury et, al., 2022 have discussed It is fast, sensitive, and reliable, which is ideal for small wearable sensors used in real-time monitoring of respiration rate. Saeed et, al., 2021 have discussed the system is the LPortable UWB RADAR sensing system, which can sense and convert small chest movements into micro-Doppler signatures using modern technologies. It further utilizes a ResNet algorithm to estimate respiratory rate with high accuracy, providing individuals with a convenient, minimally invasive way to monitor their respiratory health on the go. Lyra et, al., 2021 have discussed the camera approach, powered by deep learning, applies AI to analyze thermography images of ICU patients for vital signs monitoring. This next-generation technology has the potential to provide accurate, instantaneous feedback on heart rate, respiratory rate, and other vital signs, thereby enhancing patient care and providing early alerts to signs of critical conditions. Wu et, al., 2021 have discussed the description was as follows: The AECOPD prediction system is an AECOPD risk prediction system implemented in software, utilizing data from wearable devices and machine and deep learning models for the early recognition of those at risk for an AECOPD. It was conducted in the form of a hospital-based cohort for better screening and management of AECOPD at an early stage. Goh et al., 2021 have discussed This system accumulates and processes data in real time, providing accurate and timely information on air hygiene to inform measures that improve the air environment for passengers. Li et, al., 2023 have discussed the novel integrated wearable intelligent sensor system enables continuous real-time monitoring of several respiratory parameters. It can achieve non-invasive and remote respiratory health monitoring, offering timely and accurate information for the early prevention and intervention of respiratory diseases. The system is small in size, comfortable, comfortable, and easy to operate, making it suitable for long-term monitoring applications. Mo et, al., 2022 have discussed the cooperative three-layer architecture for non-invasive respiratory rate monitoring. This structure is a multi-level one, which employs target tracking and false peak-eliminating methods. It permits an exact and effective respiratory rate verification without direct contact. It requires the partnership of sensors, processing, and applications to achieve accurate and real-time measurement of breathing rate. Khan et, al., 2022 have discussed We develop an IR-UWB radar-based robust heart rate detection system that utilizes a deep learning method to provide an accurate measurement of a person's heart rate while in a moving vehicle. It is based on UWB radar technology to sense heart signals and utilizes deep learning algorithms to eliminate noise and motion artifacts, thereby achieving more accurate detection. Table 1: Demonstrates the Full Analysis (Menaka et al., 2022; Boopathy et al., 2024).

Table 1: Comprehensive Analysis

Author	Year	Advantage	Limitation
He et al., 2022	2022	Continuous and accurate tracking of respiratory information in real-time, facilitating prompt and precise detection of any abnormal breathing patterns or respiratory disorders.	One limitation is that the system may be affected by environmental factors such as distance, occlusion, and lighting conditions.
Alharbi et al., 2021	2021	Real-time system prediction allows for immediate detection of anomalies or potential health issues, providing timely interventions and improving patient outcomes.	Uncertainty in prediction due to constantly changing inputs and potential delays in stream processing platforms.
Bokka et al., 2021	2021	Deep learning enables faster and more accurate classification of respiration signals, allowing for real-time monitoring and potentially improving diagnoses and treatment plans.	The limitation is that the technique is dependent on high-quality input data, making it vulnerable to noise and interference.
Filosa et al., 2022	2022	"Meta-learning allows for adaptive and efficient model selection for accurate and robust respiratory flow predictions from FBG-based wearables in a variety of unrestrained environments."	The algorithm may not be able to accurately predict respiratory flow in sudden and unexpected situations or events.
Wang et al., 2021	2021	ELC model helps quantify severity of heart and respiratory issues, allowing for prompt intervention to prevent complications.	"ELC model does not take into account individual variations and may not accurately reflect changes in heart and respiratory rate."
Dai et al., 2021	2021	Real-time and continuous monitoring of respiratory rate can help detect early signs of respiratory illness or distress, allowing for prompt intervention and improved outcomes.	Reliability may be affected by skin color, motion artifacts, or placement of the device on the wrist.
Bandopadhaya, et al., 2024	2024	Early detection of silent hypoxia allows for prompt treatment and prevention of severe COVID-19 complications, potentially saving lives.	One limitation is potentially inaccurate results due to variations in patient data and demographic differences that can affect the AI model's performance.
Chowdhury et al., 2022	2022	Efficient and fast processing speeds on low-powered devices, reducing hardware and power costs for implementation.	Limited accuracy due to potential differences in data collection and need for large and diverse datasets for training.
Saeed et al., 2021	2021	Improved accuracy in detecting respiratory rate due to the use of advanced machine learning algorithms like ResNet.	Limited accuracy due to interference from external factors such as body movement, ambient noise, and clutter in the environment.
Lyra et al., 2021	2021	High accuracy and stability in vital sign monitoring due to the ability to detect subtle changes in thermography images.	The deep learning algorithm may not be able to accurately detect vital signs in patients with abnormal skin temperatures or conditions.
Wu et al., 2021	2021	Early detection of exacerbations leading to timely and effective treatment, reducing hospitalizations and improving patient outcomes.	Difficulty in accurately predicting complex medical conditions due to the potential for multiple factors and individual variability.

Goh et al., 2021	2021	The machine learning algorithm allows for accurate and near instant measurements of in-vehicle air quality, providing real-time information for immediate action.	Limited accuracy due to factors affecting air quality that are not accounted for in the prediction algorithm.
Li et al., 2023	2023	The advantage of integrated wearable smart sensor system is real-time monitoring of multiple parameters, allowing for more comprehensive and accurate data collection.	One limitation is that the system may not accurately measure respiration if the sensors are not properly placed on the body.
Mo et al., 2022	2022	Continuous and noninvasive monitoring allows for real-time tracking and detection of respiratory rate changes, improving early intervention and patient outcomes.	One limitation may be that the target tracking and false peaks eliminating algorithms may not be accurate for all types of body movements.
Khan et al., 2022	2022	Accurate heart rate detection can be achieved quickly and reliably in a moving vehicle, improving safety and monitoring capabilities.	"Limited accuracy due to environmental factors such as vehicle vibrations, signal interference, and body movements affecting signal quality."

- Accuracy and Reliability: There is a strong demand to achieve high accuracy and reliability in the detection of respiration rate, which is a real challenge. This requires a robust algorithm to accurately differentiate between respiratory and other types of motions or noises. For instance, speaking, coughing, or body movement might block respiratory monitoring and affect the results.
- Variability in Respiratory Rate: There is a considerable difference in breathing rate among individuals, or even when measured on the same individual, considering factors such as age, level of physical activity, health status, and others. Therefore, any algorithm designed to measure respiration rate that focuses solely on respiration rate needs to be adaptive and properly account for patterns in respiration.
- Time and resources: Moreover, processing load to estimate the respiration rate can add technical complexity to the above. It is desired to have good runtime complexity, space complexity, and energy consumption. This is important for real-time scenarios, such as patient monitoring in healthcare.
- Environment: Poor algorithm performance due to environmental noise. This is particularly exaggerated in the wild, where users can be taken out of the safe lab environment, and the algorithm must generalize across various conditions.

Two of today's fastest-moving technologies are deep Learning and real-time sensor data, both of which are transforming the way we handle streams of data across a wide range of applications. Photo by Markus Winkler on Unsplash. Deep Learning, a subset of Machine Learning, is AI that can learn from data; a lot of it, in this case, lots and lots of data, to perform certain complicated tasks, such as Image Recognition, Speech Processing, and Natural Language Understanding, among others. Real-Time Sensor Data: The name already implies that it's the data being collected and processed in real-time by the use of sensors in devices and machinery. These two technologies are combined, allowing deep learning algorithms to be applied to real-time sensor data for predictive and reactive insights. This cocktail facilitates faster and more accurate decision-making, making it easier to learn and adapt as the data changes. Additionally, machine learning and deep neural networks applied to real-time sensor data processing can enable the detection of complex trends and abnormalities that would otherwise be difficult to capture. Particularly in self-driving cars and predictive maintenance scenarios, environmental monitoring is crucial. It is indeed a game changer; Deep Learning and Real-Time Sensor Data potentially

have the power to revolutionize various industry segments by venturing into new levels of efficiency, accuracy, and complexity in the world of data interpretation as a whole (Zigui et al., 2024; Wu et al., 2023).

3 Proposed System

1) Construction Diagram

- **Raw PPG Datasets**

Raw PPG signals and individual PPG recordings act as vital signs and can be used for various purposes in medical and research domains. PPG is a non-invasive method of measuring blood flow in the skin using light sensors. The sensors shine a light on the skin and detect how much light is reflected from blood vessels, matching the ebb and flow of blood volume with each heartbeat. The first step involved when working with raw PPG signals is to acquire them. This is achieved by placing a PPG (photoplethysmography) sensor on the skin, typically on the finger, wrist, or earlobe.

The input signal is received by the intermediate layer, which strengthens a weighted link. The output of the natural network is given by,

$$n_i = d\left(net_i\right), \quad net_i = \sum_j \varphi_{ji} m_j + \phi_j \tag{1}$$

The bias variable is where? The frequency response of the output and the input is dictated by the frequency function and ranges from –0.1 to 0.1.

Output data control, such as managing an output layer in a deep-belief neural network, is reduced. The sensor, in turn, measures the variations in light intensity associated with pulsatile blood flow variations.

$$P = \sum_k P_k, \quad P = \frac{1}{2}\sum_k \sum_j \left(em_{kj} - m_{kj}\right)^2 \tag{2}$$

This information is subsequently digitized and recorded for later analysis. The PPG waveform, reflecting the variation of blood volume with time, provides important information on a person's cardiovascular condition.

- **Data Processing**

Data processing is the application of a process to real-life data so that such data yields a quantity. These operations are conducted on a global scale, including collection, organization, retrieval, analysis, and storage. This is crucial in today's business world, as a significant amount of data is generated daily and needs to be stored and managed effectively. First, it acquires data, which is the first step of data processing. That means pulling the "raw" data from various sources, such as databases, forms, and sensors, or receiving them manually. This data can also come in multiple types, such as numbers, text, images, or even sound. However, once data is collected, it should be formatted in a way that makes it easier to organize and analyze. Here is where structuring the information is crucial. Data is juxtaposed,

systematized, and labeled in such a way that these can not only be referred to but also manipulated easily. Data retrieval: The act of getting information from a database or other storage system. It allows the data to be converted into the information they require for analysis or decision-making.

- **Respiratory Signal Extraction**

Respiratory Signal Extraction Respiratory Signal Extraction is the process of (Extracting & Analyzing) Respiratory signals from different sources, such as Audio Recordings, Physiological signals in medical devices, Video Recordings, etc. It is a vital process in respiratory monitoring and diagnosis, allowing for the accurate monitoring of respiratory parameters and the identification of anomalies in the breathing pattern. Figure 1: Shows the construction diagram.



Figure 1: Construction Diagram

As we work on both the raw and processed signals in parallel (i.e., we extract what we wish for), the first step towards extraction that we engage in is sensor detection. Manual ROI picking within a video - Auto probing using algorithms capable of scanning audible respiratory sounds or physiological signals. After the source is detected, the signal will be pre-processed to remove any noise or artifacts that are irrelevant to the analysis. Filtering may refer to the removal of high-frequency noise or baseline drift caused by movement artifacts. These respiratory cycles consist of consecutive inspiration and expiration cycles at a frequency divided by the sample rate.

- **Respiratory Signal Labeling**

Respiratory signal labeling is determined by the process of identifying and distinguishing various patterns in the respiratory signals. This information is typically sensed and monitored, representing in time the oscillations of the breathing pattern and the expansion of the chest. Seed labeling is a process of respiratory system analysis that allows for marking and identifying various phases of the respiratory cycle (e.g., inhalation, exhalation, and pause). Acquisition of respiratory signal information: The first step in the respiratory signal processing pipeline is the acquisition of respiratory signal information. A

preferred method is to use a standard approach, such as a chest band with a pressure sensor or respiratory belts to measure changes in chest circumference (Mohameed et al., 2025).

The room parameter is obtained by differentiating P with respect to.

$$\frac{\partial P}{\partial \phi_j} = \sum_k \delta_{kj},$$
$$\delta_{kj} = \frac{\partial P_k}{\partial m_{kj}} \frac{\partial m_{kj}}{\partial net_{kj}} \quad (3)$$

Then,

$$\frac{\partial P_k}{\partial m_{kj}} = -\sum_s \delta_{ks} \phi_{sj} \quad (4)$$

When the signal is received, it is then processed to eliminate noise and any undesired features. The subsequent respiratory signal labeling task is the cycle for respiration, as described in Section I. This is usually achieved by recognizing the peaks and valleys in the signal, respectively representing in- and ex-respiration, and marking those points.

- **Train Model**

So, a Train Model is nothing more than a computer job running to build a model from your data. Model train: The train Model analyzes and finds the patterns or nature of the data so that it can predict it well from the new dataset. Choose the Algorithm in the Train Model according to the task. In the Train Model step, select the appropriate Algorithm for your task. This may be a type of information, an aim, or a willingness to receive an interpretation from a model. For example, I can tell you that an LR model that is great with a target that is a continuous numeric number is going to suck with a categorical target. In contrast, the opposite is true for decision trees (at least for all the models I've seen). Now that we have our Algorithm, we need to prepare our data for training. This encompasses all data-related tasks, including data cleaning, normalization, and feature engineering. Data cleaning involves removing noise or irrelevant information from the data while normalizing ensures that all features are on the same scale.

- **Test Model and Output Processing**

Test Model: A model used to test software, validating Software is working or correctly functionality and performance- A set of input, guidelines or instructions that describe how to test a program and what results are expected. The goal of the Test Model is to verify the program requirements and ensure the model's intent is accurately represented. There are four main phases in the life of the Test. The life of the Test Model can be divided into (a) Planning, (b) Designing "(c) Executing "and (d) Reporting. The Test specifies what is to be tested, as the testing team sets targets during the planning process, and also outlines the general test objects and the test plan itself. This plan will determine how you are going to test, which test cases are required, and the resources needed for testing. Design phase: - Actual test MODEL GETS developed. It includes the type of tests and the test case from which the tests are to be performed. The Test Model also contains the test environment and the test data required for each Test. It's the most important stage, as you're building the structure of your work before executing it.

- **Performance Evaluation**

Refers to the process of evaluating an individual or group's performance. It's usually conducted to assess the level of accomplishment relative to the goals, objectives, and targets that the organization has set. It is an invaluable instrument for gauging how productive, effective, and valuable an employee is to the company. The Performance Appraisal System includes several steps, one of which is establishing clear objectives and performance expectations. It's usually the result of cooperation between an employee and a manager as they determine what the employee will accomplish within a certain period.

The Bayesian approach requires updating and remembering the category probability.

$$E\left(r_j \mid J_{1,\ldots,s-1}\right) \tag{5}$$

$$E\left(O_w = r_j \mid J_s\right) \tag{6}$$

These must be SMART goals – specific, measurable, achievable, relevant, and time-bound. When goals are set, performance can be appraised in various ways, such as through self-evaluation, peer evaluation, or 360-degree feedback. Another common practice is to maintain a list of KPIs and metrics to track the progress of the established objectives. These KPIs encompass both qualitative and quantitative aspects, setting a standard against which performance can be measured and compared.

2) Functional Working Model

- **ResNet Layer**

The Residual Network (ResNet) layer is the fundamental building block of the ResNet architecture, a deep Convolutional Neural Network (CNN) that has been demonstrated to be effective in numerous computer vision problems. The main goal of this layer is to solve the vanishing gradient difficulty, which is the case that, as the number of layers of a deep network increase, the gradient of the loss function goes down drastically. The learning process may be hindered, resulting in low performance, particularly in accurately calculating 3D spatial information, object localization, and human motion localization. Currently, the results are rendered concerning the estimated camera gesture. The gesture of the current camera and H is the observation function.

$$E\left(r_j \mid J_{1,\ldots,s-1}\right) = \frac{1}{V} E\left(r_j \mid J_{1,\ldots,s-1}\right) E\left(O_w = r_j \mid J_s\right) \tag{7}$$

Where and denote different iterations on the training samples. The B2 model represents the training sample at the modal end with less than 1 at the modal peak with greater than 1.

$$v_s = g\left(m_s\right) \tag{8}$$

The Reset layer works by including a shortcut connection (also known as a skip connection or identity mapping) that connects the input of a layer to its output. It enables efficient gradient flow directly from input to production without passing through multiple layers. It solves the problem of disappearing gradients, where the gradient doesn't get reduced as it goes deeper through the network.

- **Remote Feature Exploration Layer**

RFEL (Remote Feature Exploration Layer) is a software layer that allows exploration and access to remote features. It is an essential component in current networked systems, particularly in the era of the

Internet of Things (IoT) and/or wireless communication networks. The latter is intended to enable a central network node to communicate efficiently and effectively with remote features located at some significant distance. Figure 2: Shows the functional block diagram.



Figure 2: Functional Block Diagram

The RFEL is mainly used for remote feature finding, matching, and tracking. These features may include sensors, actuators, and other network devices located in remote or hard-to-reach areas. This is the layer between your control system and the remote (such as a switch); it handles all communication and data exchange. RFEL operations can be divided into three primary functions: discovery, management, and communication. The first of the operations is discovery, which involves the RFEL acting as a seeker within its particular network for foreign features.

- **Downsample, Segmentation Head, Upsample, Conv 3x3, Relu, Concatentation**

It is a commonly used building block in many computer vision tasks, such as object detection and image segmentation, because it compresses an input image to a smaller one through some convolution with a kernel or filter. The segmentation head is a common component in deep learning networks for semantic segmentation. Primarily to extract features from a downsampled input image, as opposed to down sampling where the feature map from the segmentation head is stretched to be of the same size as the input image (Alzaidi, 2024).

Let us suppose that the camera device is now looking at the current gesture with z k∧. It is necessary to identify an error in this part, as a mistake could lead to conflicting observations.

$$p = \hat{v}_s - v_s \tag{9}$$

$$\min_{m} \sum_{s=1}^{y} p\left(m_s, v_s\right)^{Q} \Omega p\left(m_s, v_s\right) \tag{10}$$

$$X^{'} = floor\left(255 * \frac{X - w_{\min}}{w_{\max} - w_{\min}} \right) \quad (11)$$

Convolution 3x3 Rely stands for the convolution operation with a kernel size of 3x3, followed by an activation function similar to Relu (rectified linear), etc. Convolution is the basic operation in computer vision, where we move a filter over the input image to get the most relevant derivative features. Concatenation operates through a chaining feature map from various layers. In deep learning (DL) models, concatenation is typically used for feature fusion from one layer to the next, allowing the model to learn both low level and high-level features simultaneously.

3) Operating Principle

- **Respiratory Symptoms**

Symptoms may vary in severity across all age groups. The symptoms may also be present in other respiratory disorders affecting the lungs, airways, or chest wall. They are the result of multiple factors, including infection, inflammation, and airway obstruction. The Respiratory System The respiratory system supplies the body with oxygen and removes carbon dioxide.

Multimodal training samples need to be developed to enable deep learning.

$$\left\{ \left(m_j, n_j \right) \right\}_{j=1}^{Y} \quad (12)$$

It includes the upper and lower respiratory tracts. The upper respiratory tract consists of the nose, mouth, sinuses, and throat, while the lower respiratory tract consists of the trachea, bronchi, and lungs. The respiratory mucosa is composed of cells that cover the respiratory tract and secrete mucus, which traps foreign particles and microorganisms.

It's the equation of the decision plane that's needed when you go for classification.

$$h(m) = u^{Q} m + c \quad (13)$$

$$u^{Q} m + c = 0 \quad (14)$$

Inflamed mucosa can result in symptoms including congestion, a runny nose, and a sore throat. Respiratory tract infections, including the common cold, are the most common cause of respiratory symptoms.

- **In-Patients, ED admissions, Immunocompromised out-pts, Respiratory Film Array**

Timely and effective patient care is dependent on a competent healthcare system. A part of a healthcare system includes in-patient admissions, emergency department (ED) admissions, immunocompromised outpatients, and respiratory film array testing. A critical aspect of these procedures is that they must be performed with a high level of coordination to optimize the efficiency, precision, and timing of the patient's treatment. In-patient admissions are patients who require hospitalization for treatment and observation. This typically begins with a doctor's visit, during which the doctor assesses the patient's condition and determines whether hospitalization is necessary. Fig 3: Shows the operational flow diagram.



Figure 3: Operational Flow Diagram

Once you accept a decision, the patient is admitted, and there is a unit or service in the hospital that has your patient assigned to it. The ED is a facility that addresses a wide range of urgent conditions, from minor to life-threatening. When patients reach the ED, they are triaged according to the severity of their condition, and the most severe cases are treated first. This ensures that patients can be supported when necessary.

- **Out-pts with no High-risk Conditions; ED Parents not Being Admitted, Flu Season, Non- Flu Season**

Outpatient services (usually shortened to Outpatient) refer to the medical procedures and treatments patients receive and are discharged from the same day after visiting a hospital or healthcare establishment. Additionally, outs are crucial during flu season for assisting walk-ins with flu symptoms. Because the flu is highly infectious, we want to avoid placing these patients, who are at an increased risk of contracting the flu, near others who are also at an increased risk, such as the elderly or those with weakened immune systems. These patients could be seen as outpatients in dedicated waiting room spaces for other patients with similar symptoms and different treatment areas, thereby exposing fewer people to potential transmission. "Outpatients, however, still need basic health check-ups and minor treatments when the flu is not around." This is around the time many people get their flu shot or ensure they have all important physicals and check-ups taken care of, as long as they haven't already been exposed to the flu virus. This can be treated on an outpatient basis, is less expensive, and is faster than hospitalization.

- **Influenza Rapid Antigen**

Rapid tests for diagnosing influenza, also known as the flu, are widely used. Those tests identify specific proteins of the virus, known as antigens, in a patient's respiratory sample. It is a multi-step process; the first step is sample acquisition. The first step in applying the influenza rapid antigen test is to collect respiratory samples from the patient. It is typically performed by introducing a nasal or throat swab into the patient's nose or throat.

$$\gamma_j = n_j\left(u.m_j + c\right) \tag{15}$$

$$m_j = \frac{1}{\|u\|}.n_j\left(u.m_j + c\right) \tag{16}$$

The above parameters are calculated, and the convolutional neural network trains the entire connection layer itself. In CNNs, the coadaptation of the neurons can be reduced by properly tuning the ratio parameters.

The ground-truth trajectory serves as an empirical benchmark for evaluating tracking results against different methods, as it was used to estimate and calculate similarity.

$$\eta = \frac{1}{\left|(l-l')(k-k')(q-q')\right|'} \tag{17}$$

Where the point coordinates of the tracking trace are l′, k′, and q′, and those of the original trace are l, k and q.

The more accurate the recognition results (evaluated in the following manner), the better.

$$\omega = \frac{J_q}{J} \tag{18}$$

Where J is the total number of recognitions and refers to the number of correct recognitions, the recognition accuracies of the two algorithms are computed and determined.

The swab collects a sample of respiratory secretions, which can include the flu virus if the patient is infected. The solution vial samples are placed into is intended to help release any viral particles once the sample is collected. This solution generally contains a buffer, which allows it to open the virus and release its contents. The swab is then discarded, and the solution is agitated. For the actual antigen test, a test strip or card is employed.

4 Result and Discussion

The proposed model, RTSD-RRD (Real-Time Sensor Data - Respiration Rate Detection), has been compared with the current models of NIRRD (Novel Intelligent Respiration Rate Detection), DLBRR (Deep Learning-Based Respiration Rate), and RRL-SD (Respiration Rate Learning - Sensor Data).

- **Accuracy**

This means that the distance of the algorithm output to the ground truth respiration rate measurements is determined. High accuracy means that the algorithm can accurately determine the respiratory rate in real-time. Table.2 shows the comparison of Accuracy between existing and proposed models (Sumiati et al., 2025).

Table 2: Comparison of Accuracy (in %)

No. of Inputs	NIRRD	DLBRR	RRL-SD	RTSD-RRD
100	73.63	75.56	71.31	79.74
200	73.30	74.06	70.72	77.87
300	71.96	72.95	69.74	77.04
400	70.82	72.57	68.53	76.13
500	69.77	71.56	67.39	75.21



Figure 4: Computation of Accuracy

- **Speed**

The dispensation rate of the algorithm is crucial for real-time respiration rate monitoring. It must process the sensor data and produce a result in real time to enable monitoring. Table.3 shows the comparison of Speed between existing and proposed models.

Table 3: Comparison of Speed (in %)

No. of Inputs	NIRRD	DLBRR	RRL-SD	RTSD-RRD
100	84.63	88.56	87.31	93.74
200	84.30	87.06	86.72	91.87
300	82.96	85.95	85.74	91.04
400	81.82	85.57	84.53	90.13
500	80.77	84.56	83.39	89.21



Figure 5: Computation of Speed

- **Sensitivity**

The sensitivity of the algorithm is a measure of how well it detects small changes in respiration rate. The more sensitive the value, the more effectively the algorithm can recognize small changes in the respiration rate. Table.4 shows the comparison of Sensitivity between existing and proposed models.

Table 4: Comparison of Sensitivity (in %)

No. of Inputs	NIRRD	DLBRR	RRL-SD	RTSD-RRD
100	82.63	85.56	81.31	89.74
200	82.30	84.06	80.72	87.87
300	80.96	82.95	79.74	87.04
400	79.82	82.57	78.53	86.13
500	78.77	81.56	77.39	85.21



Figure 6: Computation of Sensitivity

- **Specificity**

This parameter indicates how well the algorithm discriminates between respiration rate and other physiological signals. The algorithm must be highly specific to record and report actual respiration rates accurately. Table.5 shows the comparison of Specificity between existing and proposed models.

Table 5: Comparison of Specificity (in %)

No. of Inputs	NIRRD	DLBRR	RRL-SD	RTSD-RRD
100	77.63	79.56	77.31	84.74
200	77.30	78.06	76.72	82.87
300	75.96	76.95	75.74	82.04
400	74.82	76.57	74.53	81.13
500	73.77	75.56	73.39	80.21



Figure 7: Computation of Specificity

5 Conclusion

In this work, a novel deep learning-based algorithm for detecting respiration rates from real-time sensor data is demonstrated. It not only improves accuracy but also reduces the time consumed by traditional methods. Based on the analysis, such a new algorithm can also be explored for other related healthcare Parameterization deep learning techniques. This algorithm can achieve high accuracy in discriminating respiration rates from sensor data, where manual counting cannot be a primary reliable measurement method, and human factors can potentially induce errors. It also provides a live capability, enabling real-time response and intervention during emergencies, thereby improving patient survival. Furthermore, the deep learning embedded in this algorithm can be fine-tuned and adapted to various medical fields. In closing, this next-generation algorithm has the potential to change how respiration is tracked, potentially providing patients with finer-featured and less noisy spectral resolution for health care. As R&D investment in this area increases, we will see continued progress in the field of medical technology as a result of the innovative algorithm.

References

[1] Alharbi, A., Alosaimi, W., Sahal, R., & Saleh, H. (2021). Real-Time System Prediction for Heart Rate Using Deep Learning and Stream Processing Platforms. *Complexity*, *2021*(1), 5535734. https://doi.org/10.1155/2021/5535734

[2] Alnaggar, M., Siam, A. I., Handosa, M., Medhat, T., & Rashad, M. Z. (2023). Video-based real-time monitoring for heart rate and respiration rate. *Expert Systems with Applications*, *225*, 120135. https://doi.org/10.1016/j.eswa.2023.120135

[3] Alzaidi, E. R. (2024). Optimization of deep learning models to predict lung cancer using chest X-ray images. *International Academic Journal of Science and Engineering*, *11*(1), 351-61. https://doi.org/10.9756/IAJSE/V11I1/IAJSE1140

[4] Bandopadhaya, S., & Roy, A. (2024). Early detection of silent hypoxia in COVID-19 pneumonia using deep learning and IoT. *Multimedia Tools and Applications*, *83*(8), 24527-24539. https://doi.org/10.1007/s11042-023-16473-9

[5] Bokka, N., Karhade, J., & Sahatiya, P. (2021). Deep learning enabled classification of real-time respiration signals acquired by MoSSe quantum dot-based flexible sensors. *Journal of Materials Chemistry B*, *9*(34), 6870-6880. https://doi.org/10.1039/D1TB01237A

[6] Boopathy, E. V., Shanmugasundaram, M., Vadivu, N. S., Karthikkumar, S., Diban, R., Hariharan, P., & Madhan, A. (2024). Lorawan Based Coalminers Rescue and Health Monitoring System Using Iot. *Archives for Technical Sciences*, *2*(31), 213-219. https://doi.org/10.70102/afts.2024.1631.213

[7] Chowdhury, M. H., Shuzan, M. N. I., Chowdhury, M. E., Reaz, M. B. I., Mahmud, S., Al Emadi, N., ... & Khandakar, A. (2022). Lightweight end-to-end deep learning solution for estimating the respiration rate from photoplethysmogram signal. *Bioengineering*, *9*(10), 558. https://doi.org/10.3390/bioengineering9100558

[8] Dai, R., Lu, C., Avidan, M., & Kannampallil, T. (2021, May). Respwatch: Robust measurement of respiratory rate on smartwatches with photoplethysmography. In *Proceedings of the international conference on internet-of-things design and implementation* (pp. 208-220). https://doi.org/10.1145/3450268.3453531

[9] Filosa, M., Massari, L., Ferraro, D., D'Alesio, G., D'Abbraccio, J., Aliperta, A., ... & Oddo, C. M. (2022). A meta-learning algorithm for respiratory flow prediction from FBG-based wearables in unrestrained conditions. *Artificial Intelligence in Medicine*, *130*, 102328. https://doi.org/10.1016/j.artmed.2022.102328

[10] Goh, C. C., Kamarudin, L. M., Zakaria, A., Nishizaki, H., Ramli, N., Mao, X., ... & Elham, M. F. (2021). Real-time in-vehicle air quality monitoring system using machine learning prediction algorithm. *Sensors*, *21*(15), 4956. https://doi.org/10.3390/s21154956

[11] He, S., Han, Z., Iglesias, C., Mehta, V., & Bolic, M. (2022). A real-time respiration monitoring and classification system using a depth camera and radars. *Frontiers in Physiology*, *13*, 799621. https://doi.org/10.3389/fphys.2022.799621

[12] Islam, M. R., Kabir, M. M., Mridha, M. F., Alfarhood, S., Safran, M., & Che, D. (2023). Deep learning-based IoT system for remote monitoring and early detection of health issues in real-time. *Sensors*, *23*(11), 5204. https://doi.org/10.3390/s23115204

[13] Khan, F., Azou, S., Youssef, R., Morel, P., & Radoi, E. (2022). IR-UWB radar-based robust heart rate detection using a deep learning technique intended for vehicular applications. *Electronics*, *11*(16), 2505. https://doi.org/10.3390/electronics11162505

[14] Kumar, A. K., Ritam, M., Han, L., Guo, S., & Chandra, R. (2022). Deep learning for predicting respiratory rate from biosignals. *Computers in biology and medicine*, *144*, 105338. https://doi.org/10.1016/j.compbiomed.2022.105338

[15] Li, Y., Liu, C., Zou, H., Che, L., Sun, P., Yan, J., ... & Wang, Z. L. (2023). Integrated wearable smart sensor system for real-time multi-parameter respiration health monitoring. *Cell Reports Physical Science*, *4*(1). https://doi.org/10.1016/j.xcrp.2022.101191

[16] Lyra, S., Mayer, L., Ou, L., Chen, D., Timms, P., Tay, A., ... & Hoog Antink, C. (2021). A deep learning-based camera approach for vital sign monitoring using thermography images for ICU patients. *Sensors*, *21*(4), 1495. https://doi.org/10.3390/s21041495

[17] Menaka, S. R., Gokul Raj, M., Elakiya Selvan, P., Tharani Kumar, G., & Yashika, M. (2022). A Sensor based Data Analytics for Patient Monitoring Using Data Mining. *International Academic Journal of Innovative Research*, *9*(1), 28-36. https://doi.org/10.9756/IAJIR/V9I1/IAJIR0905

[18] Mo, H., Ding, S., Yang, S., Vasilakos, A. V., & Zheng, X. (2022). Collaborative three-tier architecture noncontact respiratory rate monitoring using target tracking and false peaks eliminating algorithms. *IEEE Transactions on Instrumentation and Measurement*, *71*, 1-13. https://doi.org/10.1109/TIM.2022.3205644

[19] Mohameed, H., Muthazhagu, M., Punitha, K., Anvarovna, A. S., Rajendren, D. J., & Sachdeva, L. (2025). Interdisciplinary Information Systems for Health and Education Libraries. *Indian Journal of Information Sources and Services, 15(2), 11–15*. https://doi.org/10.51983/ijiss-2025.IJISS.15.2.02

[20] Saeed, U., Shah, S. Y., Alotaibi, A. A., Althobaiti, T., Ramzan, N., Abbasi, Q. H., & Shah, S. A. (2021). Portable UWB RADAR sensing system for transforming subtle chest movement into actionable micro-doppler signatures to extract respiratory rate exploiting ResNet algorithm. *IEEE Sensors Journal*, *21*(20), 23518-23526. https://doi.org/10.1109/JSEN.2021.3110367

[21] Selvakumar, K., Kumar, E. V., Sailesh, M., Varun, M., Allan, A., Biswajit, N., ... & Upasana, S. (2022). Realtime PPG based respiration rate estimation for remote health monitoring applications. *Biomedical Signal Processing and Control*, *77*, 103746. https://doi.org/10.1016/j.bspc.2022.103746

[22] Shinde, R. K., Alam, M. S., Park, S. G., Park, S. M., & Kim, N. (2022, February). Intelligent IoT (IIoT) device to identifying suspected COVID-19 infections using sensor fusion algorithm and real-time mask detection based on the enhanced MobileNetV2 model. In *Healthcare* (Vol. 10, No. 3, p. 454). MDPI. https://doi.org/10.3390/healthcare10030454

[23] Shu, S., Liang, H., Zhang, Y., Zhang, Y., & Yang, Z. (2022). Non-contact measurement of human respiration using an infrared thermal camera and the deep learning method. *Measurement Science and Technology*, *33*(7), 075202. https://doi.org/10.1088/1361-6501/ac5ed9

[24] Shuzan, M. N. I., Chowdhury, M. H., Hossain, M. S., Chowdhury, M. E., Reaz, M. B. I., Uddin, M. M., ... & Ali, S. H. M. (2021). A novel non-invasive estimation of respiration rate from motion corrupted photoplethysmography signal using machine learning model. *IEEE Access*, *9*, 96775-96790. https://doi.org/10.1109/ACCESS.2021.3095380

[25] Stankoski, S., Kiprijanovska, I., Mavridou, I., Nduka, C., Gjoreski, H., & Gjoreski, M. (2022). Breathing rate estimation from head-worn photoplethysmography sensor data using machine learning. *Sensors*, *22*(6), 2079. https://doi.org/10.3390/s22062079

[26] Sumiati., Firzatullah, R. M., Rosalina, V., & Karyaningsih, D. (2025). Development of CNN and YOLO Models to Detect Smoking Violations of Cadets in the Environment Politeknik Transportasi SDP Palembang. *Journal of Internet Services and Information Security, 15*(1), 107-129. https://doi.org/10.58346/JISIS.2025.I1.008

[27] Wang, Q., Zhang, Y., Chen, G., Chen, Z., & Hee, H. I. (2021). Assessment of heart rate and respiratory rate for perioperative infants based on ELC model. *IEEE Sensors Journal*, *21*(12), 13685-13694. https://doi.org/10.1109/JSEN.2021.3071882

[28] Wu, C. T., Li, G. H., Huang, C. T., Cheng, Y. C., Chen, C. H., Chien, J. Y., ... & Lai, F. (2021). Acute exacerbation of a chronic obstructive pulmonary disease prediction system using wearable device data, machine learning, and deep learning: development and cohort study. *JMIR mHealth and uHealth*, *9*(5), e22591. https://doi.org/10.2196/22591

[29] Wu, X., Liu, C., Wang, L., & Bilal, M. (2023). Internet of things-enabled real-time health monitoring system using deep learning. *Neural Computing and Applications*, *35*(20), 14565-14576. https://doi.org/10.1007/s00521-021-06440-6
[30] Zhang, C., Zhang, L., Tian, Y., Bao, B., & Li, D. (2023). A machine-learning-algorithm-assisted intelligent system for real-time wireless respiratory monitoring. *Applied Sciences*, *13*(6), 3885. https://doi.org/10.3390/app13063885
[31] Zigui, L., Caluyo, F., Hernandez, R., Sarmiento, J., & Rosales, C. A. (2024). Improving Communication Networks to Transfer Data in Real Time for Environmental Monitoring and Data Collection. *Natural and Engineering Sciences*, *9*(2), 198-212. https://doi.org/10.28978/nesciences.1569561

Authors Biography

Dr.J. Sylvia Grace is an Assistant Professor in the Department of Computer Science and Engineering at Sathyabama Institute of Science and Technology, India. Her research interests include artificial intelligence, machine learning, and software engineering. She has published several research papers in peer-reviewed journals and actively mentors students in computational and applied computer science research.

M. Ulagammai is an Associate Professor at Saveetha Engineering College, Chennai. Her academic focus is on computer science, data analytics, and software development. She is actively involved in teaching, mentoring students, and conducting research aimed at advancing practical applications of computing technologies.

Rajat Saini is affiliated with the Centre of Research Impact and Outcome at Chitkara University, Punjab. His professional expertise includes research analytics, institutional development, and academic performance assessment. He contributes to projects that enhance research visibility, innovation, and collaboration across disciplines, supporting data-driven strategies for institutional growth.

Harsimrat Kandhari is associated with the Chitkara Centre for Research and Development, Chitkara University, Himachal Pradesh. Her research interests include academic development, research strategy, and outcome-based institutional evaluation. She actively supports interdisciplinary research initiatives, promotes innovation in higher education, and contributes to improving institutional research quality

R. Rajalakshmi is a faculty member in the Department of Electronics and Communication Engineering at Panimalar Engineering College, Chennai. Her research expertise includes signal processing, embedded systems, and communication technologies. She is engaged in teaching, mentoring, and research initiatives aimed at developing innovative solutions in electronics and communication engineering

Dr.M.K. Prathiba is an Associate Professor in the Department of Electronics and Communication Engineering at ATME College of Engineering, Mysuru. Her research interests include communication systems, VLSI design, and embedded electronics. She has authored multiple publications and is actively involved in mentoring students and guiding research projects in advanced electronics and communication technologies.