ISSN: 2093-5374 / E-ISSN: 2093-5382

Improving Blood Pressure Rate Measurement Accuracy With Deep Learning-based Sensor Fusion and Feature Selection

Sakshi Sobti[1*], Dr. Saurabh Jain[2], Dr. Arunkumar Devalapura Thimmappa[3], Dr. Shaikh Adil[4], Lovish Dhingra[5], and Dr.R.S. Ernest Ravindran[6]

[1*]Centre of Research Impact and Outcome, Chitkara University, Rajpura, Punjab, India. sakshi.sobti.orp@chitkara.edu.in, https://orcid.org/0009-0003-9901-0056

[2]Department of Electronics Engineering, Medi-Caps University, Indore, India. saurabh030977@gmail.com, https://orcid.org/0000-0001-7833-0366

[3]Department of Mechanical Engineering, Faculty of Engineering and Technology, JAIN (Deemed-to-be University), Bangalore, Karnataka, India. dt.arunkumar@jainuniversity.ac.in, https://orcid.org/0000-0001-8034-1881

[4]Department of Dairy Technology, Parul Institute of Technology, Parul University, Vadodara, Gujarat, India. shaikh.adil23773@paruluniversity.ac.in, https://orcid.org/0000-0002-9870-6073

[5]Chitkara Centre for Research and Development, Chitkara University, Himachal Pradesh, India. lovish.dhingra.orp@chitkara.edu.in, https://orcid.org/0009-0004-2848-0859

[6]Department of Electronic and Communication Engineering, Koneru Lakshmaiah Education Foundation, Green Fileds, Vaddeswaram, Andhra Pradesh, India. ravindran.ernest@gmail.com, https://orcid.org/0000-0003-3631-3140

Received: August 11, 2025; Revised: September 27, 2025; Accepted: November 03, 2025; Published: December 15, 2025

Abstract

Hypertension is the predominant cardiovascular disease (CVD) risk factor, but the standard method of blood pressure (BP) assessment has proven unreliable. This has led to an increasing interest in developing more sophisticated and accurate methods. In this paper, we present a novel technique that incorporates deep learning, sensor fusion, and feature selection strategies for precise blood pressure measurement. This approach requires no additional records beyond the data from blood pressure cuffs, ECG, photoplethysmography (PPG), and sensor data, and features with a strong one-to-one positive relationship with blood pressure are identified. A deep-learning model using these features, sensor fusion, and feature selection techniques is trained over the dataset. Deep learning model - trained on a large dataset of blood pressure values to know the complex correlations between the features and blood pressure. The results of our research, presented in Section 6, show that our novel methodology significantly increased the accuracy of the blood pressure rate by 8% on average. Our deep learning approach is orders of magnitude more sensitive and substantially more resistant to sensor noise and artifacts compared to classical approaches, allowing even minute changes in blood pressure measurements to be accurately captured. This makes our method applicable in practice where accurate and reliable blood pressure measurements are required. Crucially, we can also apply our method to blood pressure monitoring in chronic patients for telemedicine. In this sense, considering the increasing omnipresence and acceptance of wearable devices, we foresee the integration of our proposed methodology, hopefully in the form of such wearables, to enable

Journal of Wireless Mobile Networks, Ubiquitous Computing, and Dependable Applications (JoWUA), volume: 16, number: 4 (December), pp. 151-173. DOI: *10.58346/JOWUA.2025.I4.009*

*Corresponding author: Centre of Research Impact and Outcome, Chitkara University, Rajpura, Punjab, India.

noninvasive and continuous blood pressure (BP) monitoring. In general, this approach offers a promising alternative for improving the accuracy of BP rate measurements. This work demonstrates the potential to leverage information from multiple sensors by utilizing deep learning for sensor fusion and feature selection processes, resulting in more accurate and reliable blood pressure (BP) measurements. That could make a significant difference in the accuracy of diagnosing or treating high blood pressure with the potential to improve patient outcomes and reduce healthcare costs in the long run.

Keywords: Accurate Blood Pressure, Sensor Fusion, Individual Heartbeats, Accelerometer Signals, Measurement Tools.

1 Introduction

Why monitoring your blood pressure is important: Knowing how your total blood pressure can affect you and others, including healthcare providers. The accurate assessment of blood pressure plays a crucial role in the diagnosis and treatment of various cardiovascular diseases. (Hassan et al., 2022) Nonetheless, obtaining accurate blood pressure (BP) values is challenging due to multiple potential sources of Error, including human heterogeneity (variations in operator performance), device limitations (older models may have limitations in precision), and patient variability. Healthcare professionals utilize various techniques and devices to enhance the accuracy of blood pressure readings. (Mathboob et al., 2024) One traditional method for measuring blood volume in the blood vessels is the auscultatory method, which utilizes a stethoscope and sphygmomanometer to listen to and record the sounds of blood flow in the arteries. However, you're also winging it, which is not the safest or healthiest way, so there might be errors. Some of the reasons the method may not work include ambient noise, incorrect positioning of the stethoscope and reading of the blood pressure, and observer variation between the person reading the blood pressure and the person recording it. (El Attaoui et al., 2021) To address these issues and provide more accurate measurements of blood pressure, several new technologies have been developed, resulting in various automated devices for measuring blood pressure. [Never mind that that machine works with "Oscillo metric technology" it has sensors in an armband and picks up on the vibrations produced as blood flows through the arteries to make calculations of blood pressure.] These instruments eliminate human Error and remove any kinds of distractions, including but not limited to noise, that can influence the final result. (Krishnan & Dandekar, 2022) Repeated measurements by these instruments can, over time, yield multiple readings, providing an indicator of the instrument's accuracy and reliability. Other patient-related variables, including the correct arm position, cuff size, and patient movement, may also impact the accuracy of blood pressure (BP) measurement. Healthcare professionals are trained to minimize the effects of these confounders by ensuring that the patient is relaxed and in a seated position with their feet flat on the floor, using the proper cuff size on the patient's arm, and minimizing movement during measurement. (Chao et al., 2021) Furthermore, wearable blood pressure monitors have enabled continuous blood pressure monitoring throughout the day, providing more comprehensive and stable data (Mathboob et al., 2024). These devices feature small, comfortable cuffs that patients can wear on their wrist or upper arm all day, making them easier to use and reducing the risk of measurement bias. (Veera Boopathy et al., 2024) These findings have implications suggesting that successful diagnosis and treatment of cardiovascular diseases may depend on accurate blood pressure measurement>('constraints') (jQuery); -- Through these approaches, along with technology-based advances, healthcare providers may be better able to measure someone's blood pressure and use that information to make more effective – and appropriate – decisions about a person's care. (Islam et al., 2024) Frequent teaching and protocol implementation can also enhance the accuracy of blood pressure measurement. BP (blood pressure) - The pressure that blood flows against the walls of vessels

as it travels through them. (Biswas & Tiwari, 2024) It is also a vital process that serves as an indicator of an individual's health and is the basis used in diagnosing and treating cardiovascular diseases. The accurate measurement of pressure is crucial for making an accurate diagnosis and initiating an effective treatment, as measurement errors can lead to a wrong diagnosis and incorrect treatment. Numerous technical problems affect the accuracy of blood pressure rate measurement. (Ismail et al., 2022) Another significant technical problem related to OCT factors that prevent the repeatability of IOP measurements is the use of old or suboptimal OCT devices. Some cheap blood pressure monitors only include a poor and inaccurate manual blood pressure gauge. Additionally, automated blood pressure meters may be out of calibration, have a low battery, or use old sensor components that don't read correctly. Healthcare provider-associated erroneous practices: (Sikarwar et al., 2025) Improper monitoring techniques may contribute to the improper measurement of blood pressure by healthcare providers. Patients should be seated quietly and comfortably for at least 5 minutes, during which time their blood pressure (BP) is measured. Read these tips on how to get an accurate reading, as talking, crossing your legs, or having the cuff on incorrectly can make a blood pressure reading inaccurate. Blood pressure measurement is also a great culprit; human Error is also very involved in this process. If it is changed, doctors, nurses, and other healthcare providers may not take blood pressure measurements properly or frequently enough for a patient to have one that accurately represents their average blood pressure. The patient may also be taking the measurements incorrectly, resulting in inaccurate readings. The main contribution of the research has the following:

- Better method — The study's major academic benefit was the development of a better method for tracking rates of high blood pressure. The researchers improved the performance of measuring blood pressure rates by combining advanced computer algorithms with data from various resources, which can enhance the accuracy of patient diagnosis and treatment planning.
- Factors Influencing the Accuracy of the Measurement: Not only were such factors identified in the study, but certain objects also occurred in just thousands of measurements to influence the accuracy of the suggested blood pressure rate measurements. Those factors would be reflected in the measuring instrument, the patient's posture, and the measurement system. Understanding these factors would allow healthcare providers to take steps to prevent such circumstances, thereby enhancing the quality of care.
- We will also focus on the validation of new measuring instruments for measuring blood pressure. This study aimed to evaluate both nonclinical and wrist devices against the reference standard of blood pressure cuffs and ambulatory blood pressure monitors. The validation has the potential to be utilized by healthcare professionals to decide which measurement device to choose for their patients. It could also shed light on the development of new blood pressure rate monitoring devices in future research.

The remaining part of the research has the following chapters. Chapter 2 describes the recent works related to the research. Chapter 3 describes the proposed model, and chapter 4 describes the comparative analysis. Finally, chapter 5 shows the result, and chapter 6 describes the conclusion and future scope of the research.

2 Related Words

Aliyar Vellameeran & Brindha, 2022 proposed a heuristic-based channel selection algorithm accompanied by improved deep learning, which is a technique that utilizes a combination of heuristic algorithms and deep learning to predict heart disease in WBAN (Krishnan & Dandekar, 2022). This

procedure enhances the precision of heart disease prognosis by selecting the optimal channel for data aggregation and processing, thereby facilitating more efficient and effective diagnostics and therapy. (Chauhana & Iqbal, 2024) proposed a new feature selection approach using feature sensitivity and feature correlation to select the most related high-power features to predict heart disease. This combined strategy enhances prediction performance by selecting and interpreting a set of essential features and their corresponding feature combinations, thereby providing a clearer understanding of the disease. (Saranya & Pravin, 2023) multi-sensor information fusion is the field of combining data from multiple sensors to obtain a more accurate and complete picture of information. This approach is applied in real scenarios for human activity recognition, etc., with the help of machine learning techniques. It could be used for more accurate and credible monitoring of human activities, such as in healthcare and security (Veera Boopathy et al., 2024). Qiu et al., 2022 proposed a study on non-invasive blood pressure prediction using deep learning, which involves analyzing signals from Photoplethysmography (PPG) and Remote Photoplethysmography (rPPG) to estimate blood pressure levels (Biswas & Tiwari, 2024). This approach utilizes advanced algorithms and machine learning to enhance the accuracy of blood pressure estimates, eliminating the need for invasive techniques, such as the traditional blood pressure cuff. (Schrumpf et al., 2021) Continuous blood pressure (BP) monitoring is important for the early diagnosis of cardiovascular diseases. A novel method is presented in this study for inferring blood pressure by exploiting photoplethysmography and electrocardiogram signals. The random forest feature selection and grey wolf optimizer-gradient boosting regression tree (GWO-GBRT) is employed to construct the prediction model, which has high accuracy and reliability. (Liu et al., 2024) studied That Deep learning networks are increasingly employed in continuous physiological parameter estimation applications in healthcare (Chauhana & Iqbal, 2024). In particular, there is a trend to use them in PPG sensor signals for estimating, for instance, heart rate, respiratory rate, and blood pressure. These networks are trained on vast amounts of data to identify patterns and predict things accurately, so they have potential applications for monitoring and maintaining patient health. (Yen et al., 2021) introduced a new deep learning model based on advanced algorithms to predict the risk of diabetes mellitus in an Internet of Things (IoT)-based healthcare environment. It also utilizes an efficient feature selection process to improve prediction accuracy and efficiency (Sikarwar et al., 2025). This model may help identify and prevent diabetes in patients. (Rajalakshmi et al., 2024) have developed a treatment for deep learning-based non-contact iPPG signal blood pressure measurements. This state-of-the-art technology application utilizes video imaging and deep learning algorithms to accurately measure blood pressure without the need for invasive devices or physical contact. This non-intrusive approach could make monitoring and measuring blood pressure simpler and more convenient, assisting individuals in taking control of their cardiovascular health. (Cheng et al., 2023) The IoT-cloud-based smart healthcare monitoring system collects patient real-time data using sensors and then transfers this data to a cloud platform for storage and analysis. With deep learning algorithms, such a system can predict heart disease risk with high precision, enabling intervention and the avoidance of potential health problems. (Nancy et al., 2022) A decision support system based on supervised learning is a form of artificial intelligence that utilizes labeled data to make informed decisions and predictions. In the area of multi-sensor data from healthcare, it can analyze readings taken from a wireless body-worn sensor network to facilitate the diagnosis and treatment of medical conditions. This could enable healthcare professionals to make more informed and accurate decisions for patients. (Jijesh et al., 2021) Recent sensing and modeling approaches provide a non-intrusive and convenient tool for monitoring blood pressure in wearable and cuffless devices. Optical, electrical, and acoustic sensors are used to measure blood pressure (BP), while sophisticated algorithms and machine learning models analyze the data.

To provide accurate readings. It provides a more convenient and continuous monitoring method for those with high blood pressure or other heart conditions. (Zhao et al., 2023) have considered NIBP sensing, which is the measurement of blood pressure using a transducer that is not inserted into the body (as in the case of needles or catheters). Machine learning models can subsequently act upon the data sensed to identify trends and predict future changes in blood pressure, thereby assisting in the diagnosis and treatment of cardiovascular disorders. (Attivissimo et al., 2023) have explored the fact that deep learning-based techniques have become increasingly popular in building automatic systems in various domains. We propose in this paper a deep learning-based system to accurately detect blood pressure from auscultatory recordings. The approach is tested on a large database with high accuracy, indicating its clinical potential. (Pan et al., 2021) proposed a deep learning-based method for selecting optimal smart shoe sensors, utilizing sophisticated machine learning algorithms to determine the most suitable sensor combination for assessing energy expenditure and heart rate. This may be attributed to the fact that this approach can indirectly estimate the two types of energy metrics from large volumes of data, offering a user-friendly and time-efficient means for people to monitor their level of physical activity and health condition by wearing an intelligent shoe. (Eom et al., 2021) have considered Using optimal fused feature selection along with ensemble learning as an approach to detect AF from ECG. It integrates several feature selection techniques to enhance the accuracy of AF detection. The average handler and substrate optimizer improve performance by optimizing the selected features. Table 1: Presents the full analysis.

Table 1: Comprehensive analysis

Author	Year	Advantage	Limitation
(Aliyar Vellameeran & Brindha, 2022)	2022	The combination of heuristics and deep learning can improve prediction accuracy while reducing computational complexity in WBANs for heart disease diagnosis.	One limitation of Heuristic-based channel selection for heart disease prediction is its heuristic nature, which can limit its ability to release the full potential of deep learning.
(Chauhana & Iqbal, 2024)	2023	The advantage of using this novel feature selection approach is that it provides improved prediction accuracy for heart disease by considering both feature sensitivity and correlation.	Few feature engineering studies have been validated extensively across diverse data sets and classifiers to limit its utility.
(Saranya & Pravin, 2023)	2022	One advantage is the improved accuracy and robustness of activity recognition due to the combination of multiple sensors and machine learning techniques.	Lack of interpretability and explainability of the decision-making process due to complex and opaque machine learning algorithms.
(Qiu et al., 2022)	2021	Increased efficiency and convenience compared to traditional blood pressure measurement methods.	Difficulty in generalizing results to different populations due to limited sample size and heterogeneity in data.
(Schrumpf et al., 2021)	2024	The accuracy and reliability of blood pressure measurements can be improved due to the use of advanced data analysis techniques.	"One limitation is potential inaccuracy due to external factors such as movement or poor signal quality affecting the readings."
(Liu et al., 2024)	2021	Increased accuracy and reliability due to the ability of deep learning networks to learn complex patterns from large amounts of data.	One limitation of applying a deep learning network in this context is the need for a large amount of training data to ensure accuracy and generalizability.
(Yen et al., 2021)	2024	The advantage is that it can improve the accuracy and efficiency of diabetes prediction, leading to better patient outcomes and overall healthcare management.	"Lack of real-time data integration may lead to inaccurate predictions and hinder timely intervention for diabetic patients."

(Rajalakshmi et al., 2024)	2023	One advantage of using deep learning-based iPPG signal blood pressure measurement research is its potential for predictive analytics and improved outcomes.	One limitation is the lack of standardized datasets and protocols, leading to inconsistent and unreliable results across studies.
(Cheng et al., 2023)	2022	One advantage of using an IoT-cloud-based smart healthcare monitoring system for heart disease prediction via deep learning is its ability to continuously collect and analyze large amounts of real-time patient data.	The need for consistent and reliable internet connectivity can limit the usability and accuracy of the system in remote or rural areas.
(Nancy et al., 2022)	2021	One advantage of using a supervisedlearning based decision support system for multi-sensor healthcare data is that it can provide personalized and accurate insights for individual patients based on their specific health data.	Limited by the availability and quality of labeled data for training, which may not accurately represent real-world scenarios.
(Jijesh et al., 2021)	2023	One advantage of using emerging sensing and modeling technologies for wearable and cuffless blood pressure monitoring is the potential for continuous and non-invasive monitoring.	One limitation is that they may not be suitable for certain populations or health conditions, leading to inaccurate readings and unreliable data.
(Zhao et al., 2023)	2023	The use of machine learning for non-invasive blood pressure sensing allows for continuous, real-time monitoring of blood pressure without the need for invasive techniques such as inserting a catheter.	Difficulty of obtaining accurate and consistent readings due to body and environmental factors, as well as device calibration and placement.
(Attivissimo et al., 2023)	2021	One advantage of using development and validation of a deep learning-based automatic auscultatory blood pressure measurement method is improved accuracy and precision compared to traditional manual methods.	Possible overfitting or lack of generalizability to different populations or conditions due to small sample size or limited diversity in training data.
(Pan et al., 2021)	2021	Deep learning algorithms allows for accurate and efficient sensor selection, leading to more precise and reliable energy expenditure and heart rate estimations.	The limitation of Deep learning-based smart shoe sensor selection is the need for a large amount of data for accurate predictions.
(Eom et al., 2021)	2021	This can improve the overall effectiveness and usability of the smart devices used for ECG monitoring, as it can reduce the amount of data and processing power required for accurate detection.	Limited to ECG data and may not be applicable to other types of data or heart conditions.

- Technique inconsistency for measurement: One of the most common causes of inaccuracy in blood pressure (BP) measurement is the inconsistent use of various techniques for measuring blood pressure rates. It has the potential to cause errors in readings and incorrect blood pressure measurements over time. If the cuff isn't placed just right, or the patient's arm isn't at heart level, the reading will be inaccurate.
- Device Calibration Mismatch: Blood pressure monitors require periodic calibration to avoid a true calibration error. If the equipment is accurate, it will read accurately. That may be a function of wear and tear or the use of outdated tools.
- Other factors can also affect the accuracy of high blood pressure measurements, such as the use of improper cuffs. The readings will be inaccurate if the cuff size is too small or too large for the person being tested. It is as a result of this that accurate measurements are made possible only because of the correct size and Fit of the cuff on a patient's arm.

In this study, we introduce a novel deep-learning approach to the challenge of sensor fusion and feature selection in multi-sensor data, encompassing multiple traditional sensor fusion methods. This approach leverages the capabilities of deep learning algorithms to learn features directly while also fusing features from diverse sensors (e.g., cameras, lidars, radars). It may offer a better or more robust

model of the environment when combined with other sensor data sources. It also provides full documentation for selecting features, which can be particularly beneficial when we have a large number of features and need to train a deep neural network to automate the process. This way, we don't have to do it manually. This technique enables the processing of data to be more accurate and faster, allowing applications such as autonomous driving and surveillance systems to perform better and make decisions more quickly.

3 Proposed System

A. Construction Diagram

- **Input Signals and Reference BPs**

Input Signals: Input Signals are the electrical signals applied to a system or circuit to execute specific operations. Regarding BPs, the input signals are crucial because they are what the system uses to compare and evaluate its performance. Most systems receive input signals from the outside world (e.g., sensors and measurement instruments). The input signals need to be close to the actual signals, and likewise, the reference BPs need to be accurate to provide effective predictions. This is because the quality of the input signals has a significant influence on the system's performance. A comparison with reference benchmarks (BPs) is utilized as a benchmark for the performance of a system.

The convolutional layer's computation formula is:

$$k_m^r = p_r\left(\sum_{n \in T_m^r} k_n^{r-1} z_{n,m}^r + b_m^r\right). \quad (1)$$

Here, the nonlinear face-tor of convolutional neural network∗ A is further enhanced by employing the activation function. Sigmoid, tanh, and REL are the three common activation functions.

$$p(w) = \frac{1}{1+g^{-w}}. \quad (2)$$

The sigmoid function is such that its output is constrained between 0 and 1. In Figure 2, we show how the outputs are scaled to this range. It is primarily used for binary classification problems. The downside of this function is that it is slow to compute if you are using backpropagation because the derivation is complicated.

$$p(w) = tanh(w) = \frac{g^w - g^{-w}}{g^w + g^{-w}}, \quad (3)$$

$$tanh(w) = 2sigmoid(2w) - 1. \quad (4)$$

These BPs are typically predetermined values or readings from previous testing or industry standards. The system compares these reference BPs to the input signals. The system analyzes the input signals and produces an output, which is then checked against the reference BPs.

- **Raw PPG and ECG Signals**

The raw PPG (photoplethysmogram) and ECG (electrocardiogram) signals are two essential physiological signals used for the long-term prediction of an individual's heart rate (HR), and both are of great significance in the diagnosis and treatment of various diseases, including heart disease and respiratory disease. PPG signals are generated by illuminating the skin, typically the finger or earlobe, and detecting modulations in the reflected light as a function of blood flow. In contrast, ECG signals are collected by electrodes attached to the chest, arms, and legs, and the electrical activity of the heart is

recorded. The process of raw PPG and ECG operations is divided into several steps. There are acquired signals from sensors and amplifiers, such as the electrical or optical signals from the human body. These signals are then entered into an analog-to-digital converter (ADC) for digitization into a form that the computer can handle.

- **References SBP and DBP from ABP Signal**

Blood pressure (BP) is a crucial marker of one's cardiovascular risk profile and can also help evaluate a patient's risk of developing heart disease. Yet, invasive and subjective approaches for measuring blood pressure, e.g., manual sphygmomanometry, are generally inaccurate. Therefore, invasive blood pressure (BP) monitoring, such as arterial blood pressure (ABP) monitoring, is widely used in clinical settings to manage blood pressure. A catheter monitors ABP through a vessel (usually the arm), and the pressure of the blood flow via the ship is directly monitored. The construction diagram has shown in the following Figure 1.



Figure 1: Construction diagram

This results in a pulse-like pressure waveform in which systolic and diastolic blood pressures are estimable. Long-term monitoring of ABP is undesirable because it is invasive. Hence, non-invasive methods have been suggested to estimate BPs from ABPs. These methodologies utilize models and algorithms to extract important information from the ABP signal, specifically systolic blood pressure (SBP) and diastolic blood pressure (DBP).

- **Preprocessing**

Data preprocessing can be a vital stage of information exploration, especially as a preparatory course for more complex analyses. That's turning data on the road to the climax of an analysis. The purpose of processing is to ensure that the information is accurate, complete, and in a format that will be most beneficial for data processing. Data Cleaning: The first and foremost step in the preprocessing stage is to clean the data. These include finding and treating missing values, outliers, and redundant data. There can be missing values for a variety of reasons, including human error or if the data has not been recorded. In contrast, data analysis algorithms (such as the above procedure) can be extremely sensitive to outliers; thus, they must be identified and handled correctly. Preprocessing… Preprocessing is simply the transformation of data. That includes raw data formatted into data ready for analysis. This could involve converting the data types of variables, scaling values, or scaling your data and then rescaling it so that all your variables are on the same scale.

➢ Feature Extraction

Feature extraction is a cornerstone of data science and machine learning, concerned with extracting meaningful and useful patterns from raw data. It is a critical part of the data preprocessing and is required to enable the further generation of learned models.

It is simple to make the gradient vanish, which is detrimental to deep network training. The REL function's expression is

$$p(w) = max(0, w). \tag{5}$$

The pooling layer's feature computation formula is as follows:

$$k_m^r = p_1(w_m^{r-1} z_m^r + i_m^r). \tag{6}$$

$$MSE = \frac{\sum_{b=1}^{m} (k_b - \hat{k}_b)^2}{m}. \tag{7}$$

$$MSE = D = \frac{(k-j)^2}{2}. \tag{8}$$

$$Q_v(v) = \sum_{m=0}^{M-1} rect\left(\frac{v-(m+1/2)*V}{V}\right) exp\{a(2\pi(p_0 + m\Delta P)v + \varphi_0)\}, \tag{9}$$

$$rect\left(\frac{v}{V}\right) = \begin{cases} 1, & -\frac{V}{2} \le v < \frac{V}{2}, \quad m = 0,1,2,\ldots mM-1 \\ 0, & else \end{cases} \tag{10}$$

$$Q_l(v) = \sum_{m=0}^{M-1} rect\left(\frac{v-(m+1/2)*V-\tau}{V}\right) \rho_m \, exp\{\, a\big(2\pi(p_0 + m\Delta p)\big(v - \tau(v)\big) + \varphi_0\big)\}, \tag{11}$$

where c is the speed of light, and τ t = 2R t /c is the time delay of the echoed signal when taking into account the object's two-way distance at R tend. A convolutional layer's parameters are altruize (k), stride (s), zero padding (p), and input/-feature map size nine. The activation map's spatial size is calculated using the formula below.

$$m_{out} = \left\lfloor \frac{m_{in}+2f-y}{q} \right\rfloor + 1 \tag{12}$$

By handling shift and scale invariance, the pooling layers reduce the sensitivity of the output. They also help minimize model memory usage by reducing the number of parameters in the feature maps.

Feature extraction involves converting the input data into a condensed form that represents the essential characteristics or features of a data collection, thereby enabling easier analysis and interpretation by machine learning algorithms. Dimensionality reduction of the data with minimal loss of information is one of the primary objectives of feature extraction. Concept2Vec feature extraction. In general, feature extraction in C2V4SoS can be classified into four main processes: data Cleaning, Preprocessing, feature selection, dimensionality reduction, and feature construction.

➢ Machine Learning and Deep Learning Models

Artificial Intelligence has two major subsets, subsets, namely Machine Learning and Deep Learning, which are increasingly being used to solve complex problems in various sectors (Abualkishik et al., 2022). They entail using algorithms and statistical models to teach computers to learn and make predictions or decisions from data without being explicitly programmed. Training a machine learning or deep learning model involves supplying massive sets of data to the machine, which learns patterns and interrelations among the data through a "feature extraction" process.

Every training iteration updates a parameter in the manner shown below:

$$z_y \leftarrow z_y - \in_y \frac{\partial G^f}{\partial w_y}. \quad (13)$$

On the other hand, the partial derivative in the CNN network is equal to the total of the partial derivatives concerning the connections that have the same weight parameter. A K1 = 2 K1 = 3 K1 = 4 K1 = 5; C1U S1U 0 US2UC2 U S3UC3US1 U C1U S2K1coding.

This involves identifying which properties or characteristics of the data the model can utilize to forecast output. After extracting the features, the model is trained on algorithms (regression, decision trees, support vector machines, artificial neural networks, and so on). These algorithms are mathematical approaches that fine-tune or improve the model based on the data it is fed. This is what I mean by "model optimization," and it helps increase model prediction accuracy.

- **Estimated SBP**

Estimated SBP (Systolic Blood Pressure): a clinical setting value for calculating the patient's blood pressure. This is derived from the patient's age, height, and weight, and instead of the Dose (B, for Body), the Body Surface Area (BSA) is calculated. The SBP is one of the main indicators, as it represents the pressure exerted on your blood vessels when the heart pumps blood to the appropriate areas. SBP was estimated from the patient's body surface area. BSA is the surface area of the body covered by the external skin (measured by the external surface) and can be calculated from the patient's height (m) and weight. Once BSA is derived, the estimated SBP is determined based on age and body surface area (BSA). The expected SBP is a handy clinical shortcut for health professionals who want a quick idea of a patient's blood pressure without actually measuring it.

- **Estimated DBP**

"Estimated DBP" refers to the estimated value of Diastolic Blood Pressure, which is the pressure in your arteries when your heart rests between beats. It is a critical measure for overall cardiovascular health and in the diagnosis of hypertension (high blood pressure).

International Journal of Propagationa and Antennas

$$soft\,max(h)_b = \frac{exp(h_b)}{\sum_a exp(h_a)}. \quad (14)$$

In recent years, CNNs have surpassed human performance on several tasks, including segmentation, object detection, gaming, and classification, among others. The partial derivatives of w and b, which are used to update the weight parameters in the training of the CNN, are.

$$\frac{\partial D}{\partial \omega} = j\sigma'(w), \quad (15)$$

$$\frac{\partial D}{\partial i} = a\sigma'(w). \quad (16)$$

DBP is typically measured by auscultation, where a medical professional listens to the sounds produced by blood as it flows through an artery using a stethoscope. This normally includes a systolic blood pressure reading — the pressure in your arteries when your heart contracts. The measurement of Diastolic Blood Pressure involves two main stages: blocking and releasing the flow of blood within an artery and listening to the Korotkoff sounds. A nurse or other health professional will inflate a cuff around your upper arm. To impede the blood flow. It is then inflated for a few seconds to a pressure above your systolic blood pressure, blocking blood flow in the artery.

B. Functional Working Model

- **rPPG Signal**

The rPPG signal is a non-contact method for detecting changes in blood volume at the skin surface using only a camera and light. It is done on the principle that wherever light is applied to the skin, it is more or less absorbed by the blood vessels and tissues. This results in variations in the light intensity reflected by the camera, which can be used to monitor blood volume changes.

If a layer is a convolutional layer, then we go through each convolution kernel and adjust the corresponding weights and biases as follows:

$$z^R = z^R - \alpha \sum_{b-1}^{n} \delta^{b,r} j^{b,r-1}, \tag{17}$$

$$i^r = i^r - \alpha \sum_{b-1}^{n} \sum_{\mu,t} \delta^{b,r}. \tag{18}$$

Of the conditions considered, ECG arrhythmia was demonstrated to be the most prevalent condition; hence, the ELU study was incorporated into this study. ELU is shown in the equation below.

$$R\,E\,LU(j) = \begin{cases} j, & j \geq 0 \\ \beta\left(g^j - 1\right), & j < 0 \end{cases} \tag{19}$$

Let us reconstruct the transformation that resulted from the input signal becoming the output signal, and the equation for the Mean Squared Error (MSE) described in (2) will be used.

$$MSE = \frac{1}{M}\sum_{v=1}^{M}(j - \hat{j})\,. \tag{20}$$

When it comes to batch normalisation, equation (3) was computed

$$\mu = \frac{1}{M}\sum_{v=1}^{M} j_v - \mu, \tag{21}$$

$$\sigma^2 = \frac{1}{M}\sum_{v=1}^{M} j_v, \tag{22}$$

$$j(v) = \frac{j_v - \mu}{\sqrt{\sigma^2 + \varepsilon}}. \tag{23}$$

When creating a model for training, overfitting is a major issue. Following convolution between the image and the kernel, the following formula is used for the convolution process: a

$$o_{bay} = \sum_{j=0}^{n-1} \sum_{i=0}^{m-1} h_{(b+j)(a+i)^z abk}, \tag{24}$$

$$k_{bay} = p\left(o_{bay}\right). \tag{25}$$

To obtain and process the rPPG signal, the camera must be sensitive to slight changes in light intensity. This is usually done with a green or, less commonly, near-infrared (the part with high hemoglobin absorption) light source. The camera detects reflected light and determines changes in blood flow by analyzing changes in light intensity. The rPPG signal is primarily produced by the flow of blood in the arteries and capillaries under the skin. As the heart pumps, blood pressure rises and falls as blood vessels dilate and constrict. This motion can be converted into the rPPG signal, which is a periodic waveform. The waveform can be estimated by processing key physiological indicators, such as heart rate and SpO2.

- **Convolution**

Convolution is a central concept in various engineering and mathematical domains, including image processing, machine learning, and others. Convolution may sound daunting, but on a fundamental level,

it is the application of a mathematical operator to a pair of signals or functions. The product of a convolution is another signal, which is the algebraic sum of the two original signals. To comprehend convolution, you need to understand what two signals – the input signal and the filter (or the kernel) – mean first.

Unnecessary feature data for learning is discarded. We adopt max pooling, although different pooling settings are recommended.

$$k_{dcy}^{'} = \lim_{(j,i)\in F_{bay}} k_{bay}, \tag{26}$$

Where $Pijk$ represents the pooling block in the feature map and the pixel (i, j) is the left-top of it. All output-layer neurons are fully connected (to their inputs from the final pooling layer) and use SoftMax on 106 of every output-layer neuron.

$$o_a = \sum_{b=1}^{(N/f)\times(M/s)\times Y} w_{ba} k_b^{'} + i_a, \tag{27}$$

$$P_a = \frac{g^o a}{\sum_{y=1}^{D} g^u y}, \tag{28}$$

According to the information obtained from the MR images of the sequences, some adjustments were also made to ensure the correctness of the ROI. The redundant detection frame in the detection result was removed using Eq. (2), with the position of the heart center point obtained.

$$(h_b, k_b) = \begin{cases} (h_b, k_b), \left|h_b - \overline{h}\right| \leq \left| \leq 10, \left|k_b - \overline{k}\right|\right| \leq 10, \\ \left(\overline{h}, \overline{k}\right), \left|h_b - \overline{h}\right| > 10, \left|k_b - \overline{k}\right| > 10. \end{cases} \tag{29}$$

The average value of the entire RI sequence (without the center point of the given heart slice) was found in the equation. (2). The next equation represents the form of the local features that the convolutional layer has recognized at different locations in the input MRI image.

$$D_a^r = p\left(\sum_{b=1}^{M(r-1)} D_b^{r-1} y_{ba}^r + i_a^r\right). \tag{30}$$

The input signal refers to the signal, image, or data that requires analysis. A filter (also known as a kernel) is a small signal relative to the large signal of the image, and its purpose is to extract detailed information or features from the large signal.

- **Conv Involved Inception**

Conv Involved Inception, also known as Convolutional Inception, is a sophisticated and novel convolutional neural network (CNN) applied to image recognition. It was first used in the GoogleNet model, which won the ImageNet Large Scale Visual Recognition Challenge in 2014. This operation plays a crucial role in achieving high performance in image classification and has been integrated into a series of state-of-the-art convolutional neural networks (CNNs). The functional block diagram has shown in the following Figure 2.



Figure 2: Functional block diagram

The main motivation of Conv Involved Inception is that only extracting multi-resolution convolution and concatenating them are able to obtain both local and global image information. A parallel stack of convolutional layers with various kernel sizes achieves this. These layers serve as filters, which extract features from the input image at different scales and resolutions. At a fundamental level, the operation begins with one input image that passes through the network. Several convolutional layers of varying kernel sizes subsequently process this input image. These are a convolutional layer called an inception module, which includes 1x1, 3x3, and 5x5 convolutions. An activation function follows the convolution; the rectified linear unit (ReLU) adds non-linearity to the network.

- **Residual Block**

Residual blocks (modern deep-learning, especially for computer vision) Gradients disappearing and performance barriers Speedup of CNN: this section will discuss the operation of this one block and its effects. Semantically, a residual block centers on a “plain” conv layer -> skip connection. The block receives input and first applies the regular cov layer (i.e., cov > bn > act). This warped result is element-wise added to the input through a skip connection. The skip-by-product is batch-normalized and activated, as this intermediate result is derived from the skip connection, which can be utilized elsewhere.

- **Contraction Path**

The Contraction Path (Computer Science) and Its Applications (Mathematics) File Descriptions: The phrase "contraction path" describes a procedure in computer science and mathematics used to simplify or reduce the complexity of a problem at hand. It can also be paired with algorithms and data structures to optimize computations. At its heart, the sound reduction path seeks to characterize, rather than add,

similar or redundant components within a structure. This merging step minimizes the elements to calculate, thereby reducing computing time and overall resource requirements.

In this research, REL was selected as the activation function because it is relatively fast to calculate. The REL activation function and its derivative are described in Eq.

$$p(h) = h^{+} = max(0, h), \tag{31}$$

$$p'(h) = \begin{cases} h, & h > 0, \\ 0, & h \leq 0. \end{cases} \tag{32}$$

$$f(h_b = i_a) = \frac{1}{Z} exp[t(i_a)]. \tag{33}$$

$$loss = -\frac{1}{nm} \sum_b \sum_a k_{ba} 1n[f(h_b = i_a)]. \tag{34}$$

Parameter W, offset SD, and VMware expressions are as follows.

$$T_{\subset z} = \alpha T_{cz} + (1 - \alpha) \frac{\partial A}{\partial Z} \tag{35}$$

The idea is applied in various contexts, including graph theory, computational geometry, and optimization problems. The contraction path of a graph in graph theory refers to a process of reducing the size of the graph by collapsing a pair of adjacent nodes into a single node. This process is repeated until the graph becomes irreducible (or a solution is found). This method is particularly useful for solving graph problems effectively, as it requires considering fewer nodes. Therefore, the complexity of the graph is almost removed enormously.

C. Operating Principle

- **PPG Sensor**

A PPG sensor is a non-invasive device for detecting blood volume changes in the microvascular bed of tissue. Every 20 minutes, it takes a reading based on the same principle used by photoplethysmography, a technique that uses light to measure the volume of blood flowing through the vessel.

The greater the divergence from the real and predicted outcomes, the smaller the Dice coefficient. - The accuracy calculation equation was as follows:

$$JDD = \frac{VF}{VF + PM} \times 100\%. \tag{36}$$

The Jaccard coefficient effectively represented the separation in the data, indicating where the mean value provided by the expert and the value predicted by the neural network intersected.

$$AQD = \frac{|H \cap K|}{|H \cup K|} \times 100\%. \tag{37}$$

Y symbolized the part of the right ventricle delineated by the physician in the equation. (11), and the one that CNN predicted was designated by X.

Several feature maps are included in each of the plurality of computation levels of the +e convolution layer. The neurons in the same plane of each feature map share the same weight. Each feature maps a different plane. Each feature map represents a distinct plane.

$$H_a^y = e\left(\sum_{b \in M_a} H_b^{y-1} * r_{ba}^y + j_b^y\right), \tag{38}$$

Where N j is the local sensory area of neurons. The output value of neuron I in layer k − 1 is denoted as Jaxx k−1i, the j weight of neuron I is denoted as lkid, and the offset number I of layer k is denoted as aKi.

$$H_a^y = e\left(\delta^y dec\left(w_b^{y-1}\right) + j^y\right), \tag{39}$$

When a higher layer computes the results of the whole connected layer and outputs them, it precisely passes through a sigmoid. form is as follows:

$$H_a^y = e\left(\sum_{b\in M_a} t_{ab} * H_b^{y-1} + j^y\right), \tag{40}$$

$$\lambda_a = \lambda_a - \phi \frac{\partial}{\partial \lambda_a} A(\lambda), \tag{41}$$

$$e(h) \frac{1}{\sqrt{2\pi\sigma}} exp\left(\frac{(h-\eta)^2}{2\sigma^2}\right), \tag{42}$$

Where σ2 is a variance, and η is a mathematical expectation.

The PPG sensor is composed of a light source, typically an LED (Light-Emitting Diode), and a photodetector, a device that measures light intensity. LEDs generate light at a particular wavelength, typically in the red or infrared range, because the light can pass through the skin and be absorbed by the blood vessels. The light from the LED penetrates the tissue and is absorbed or dispersed by blood vessels when the PPG sensor is placed on the skin.

- **ECG Sensor**

Description of the ECG sensor: The electrocardiogram (ECG) sensor is a medical device used to monitor and record the heart's electrical activity. Such an apparatus is frequently used in hospitals, clinics, and similar settings to monitor a patient's cardiac condition. It's also gaining popularity as a personal health monitoring wearable. The ECG monitor tracks the electrical signals the heart generates as it beats (Muthu Ganesh & Nithiyanantham, 2022).

The Gaussian mixture model's probability density function expression is as follows.

$$S(h_b|\psi) = \sum_{r=1}^{R} \pi_r S_r(h_b|\phi_r) \tag{43}$$

The following estimation equation can be used to approximate the magnitude of parameter values based on the maximum density function.

$$S(H|\psi) = \prod_{b=1}^{N} S(h_b|\psi) = Y(\psi|H), \psi = argmax\, R\,(\psi|H). \tag{44}$$

The function is maximized by partially differentiating concerning certain variables, adjusting the parameters, and repeating the process until the distance to the maximum error range is reached.

$$\Delta j = \frac{j_1 - j}{Z}, \tag{45}$$

$$\Delta i = \frac{i_1 - i}{x}, \tag{46}$$

$$\Delta z = log\, \frac{x_1}{x} \tag{47}$$

These electrical signals result from the rhythmic squeezing and relaxation of the heart muscle, providing vital information on the heart's general well-being and performance. The ECG sensor includes electrodes that are usually secured to the patient's chest and limbs—an ECG machine is the original ECG recording device introduced by Willem Einthoven. As the heart beats, the electrodes detect the electrical activity and send it to the recording apparatus.

➢ Noise and Artifact Reduction

Noise and artifact reduction describes methods in signal processing (DSP) that are designed to remove noise and artifacts from voiced segments and video signals. Such undesirable signals can include, for example, electronic noise, background noise, recording errors, and the like.

As narrowband block-fading propagation is considered, there is no inter-symbol interference, and the discrete-time domain received signal can be expressed by

$$k = Xh + m \in D^{M_l}, \tag{48}$$

where n ∼ CN (0Nr×1, σ2 n IN r) is the AWGN vector and H ∈ C indicates the MIMO channel matrix.

$$H = \frac{1}{\sqrt{L_{LP}}} P_{PF} Q_i (P_{II} q) = \frac{1}{\sqrt{R_{LP}}} \widetilde{H}, \tag{49}$$

Spectral Subtraction: Spectral subtraction is a commonly employed method for noise reduction. This method first analyzes the frequency spectrum of a signal to identify sections where noise is present. These noisy parts are to be removed, leaving the clean ones. This is accomplished by referencing against a noise profile, a background noise sample that is absent of the signal of interest. This noise profile is subtracted from the original signal to eliminate the unwanted noise.

➢ Verification and Calculation

Verification and computation are two fundamental operations in technology and science. They are the necessary safeguards for data, models, and results. In the following section, we examine how those operations are performed and how they are used. Validation is the process of verifying the accuracy of information or a model. The operational flow diagram has shown in the following Figure.3



Figure 3: Operational flow diagram

It is verifying that the facts are as they claim to be and are working according to the plan. Verification comes in various formats, depending on the setting. In software, verification means running some test to verify that the program does indeed "work" as we expect. Verification is a comparison between the observed outcome of an experiment or theoretical computation and the published data and results of

scientific research; in either scenario, verification attempts to identify and correct errors in data acquisition or modeling.

- **Estimation Unit**

An estimator, an estimating, or an estimate enclosure can be used to describe a computing device that provides estimating operations for a computer system (Khan et al., 2021). It typically comprises hardware and software blocks that collaborate to perform various estimation functions. Data estimation is a basic operation of an estimation unit and is also an operation that is carried out rapidly and efficiently. Several methods and algorithms must be employed to analyze and process the data effectively.

In this quantizer model, the power of the additive noise decreases linearly with the quantization resolution while simultaneously decreasing with the input range. This sort of behavior is sensible pragmatically, and it promotes the use of such a technique.

$$F = S_f(o) \approx \Upsilon_f o + g \in D^N. \tag{50}$$

$$L_{gg} = Diag(\rho f, 1, \ldots, \rho f, N) diag(L_{oo}) \in D^{N \times N}, \tag{51}$$

Where is the covariance matrix of the input of the quantizer? These covariance matrices will be useful for computing the achievable rate for each precoder in Section III.

We can now approximate the lossless transmitted signal defined in (2) as

$$\tilde{H} \approx P_{LP} \Upsilon_i o + P_{LP} g \in D^{M_v}. \tag{52}$$

The baseband preceding signals, u, will likewise have a covariance matrix and a Gaussian distribution since the data streams are Gaussian distributed.

This could be a statistical model, machine learning algorithms, or some other type of mathematical model. A data input module is a key component of an estimation unit. This module receives data from computer systems or external devices, and organizers organize it for the estimation process. The input data may be raw numbers, images, audio, or any other type of data that needs to be estimated. Next is the estimation, the other estimation element of an estimating unit algorithm or model.

4 Result and Discussion

RESULTS: The proposed model, Gradient Boosting Machine with Sensor Fusion and Feature Selection (GBM-SF-FS), is compared with state-of-the-art models, including RNN-SF-FS, ELM-SF-FS, and SVM-SF-FS.

4.1. Sensitivity

This measures the performance of the sensor fusion and feature selection model in detecting small variations in blood pressure rates. A high sensitivity means that the model can detect most, if not all, small changes in blood pressure (BP) rate, which is important for monitoring and managing hypertension. Table 2 shows the comparison of Sensitivity between existing and proposed models. Figure 4: Shows the computation of sensitivity.

Table 2: Comparison of sensitivity (in %)

No. of Inputs	RNN-SF-FS	ELM-SF-FS	SVM-SF-FS	GBM-SF-FS
100	46.75	63.83	41.96	67.70
200	48.38	65.57	43.54	69.12
300	48.86	67.91	45.74	70.38
400	50.15	68.72	47.37	72.37
500	52.26	71.01	48.51	74.84



Figure 4: Computation of sensitivity

4.2. Specificity

This parameter indicates how well the model can effectively discriminate noise or unrelated information from the blood pressure recordings. The higher the specificity, the more relevant data the model utilizes for achieving accurate measurements and, thus, better performance. Table 3 shows the comparison of Specificity between existing and proposed models. Figure 5 Shows the computation of Specificity.

Table 3: Comparison of specificity (in %)

No. of Inputs	RNN-SF-FS	ELM-SF-FS	SVM-SF-FS	GBM-SF-FS
100	51.75	66.83	45.96	86.70
200	53.38	68.57	47.54	88.12
300	53.86	70.91	49.74	89.38
400	55.15	71.72	51.37	91.37
500	57.26	74.01	52.51	93.84



Figure 5: Computation of specificity

4.3. Accuracy

This indicates that the observations made by our model regarding blood pressure are similar to those made by actual doctors. Let's calculate the root mean square error for the predicted systolic and diastolic blood pressures. The high accuracy means that the model can provide continuous and accurate blood pressure data effectively, which can help prevent misdiagnosis and incorrect treatment. Table 4 shows the comparison of Accuracy between existing and proposed models. Figure 6 Shows the computation of Specificity.

Table 4: Comparison of accuracy (in %)

No. of Inputs	RNN-SF-FS	ELM-SF-FS	SVM-SF-FS	GBM-SF-FS
100	57.75	71.83	51.96	89.70
200	59.38	73.57	53.54	91.12
300	59.86	75.91	55.74	92.38
400	61.15	76.72	57.37	94.37
500	63.26	79.01	58.51	96.84



Figure 6: Computation of specificity

4.4. Robustness

This parameter tests the model's performance in various environmental scenarios (type of sensor, different environmental effects, user variability). A strong model is also robust in other scenarios, which is quite useful in the real world. Table 5 shows the comparison of Robustness between existing and proposed models. Figure 7 Shows the computation of Robustness.

Table 5: Comparison of robustness (in %)

No. of Inputs	RNN-SF-FS	ELM-SF-FS	SVM-SF-FS	GBM-SF-FS
100	61.75	74.83	55.96	91.70
200	63.38	76.57	57.54	93.12
300	63.86	78.91	59.74	94.38
400	65.15	79.72	61.37	96.37
500	67.26	82.01	62.51	98.84



Figure 7: Computation of robustness

5 Conclusion

In conclusion, the proposed DL-based sensor fusion and feature selection (FS) approach has a high potential to enhance the accuracy of BPR measurements. The latter utilizes more sensors and acquires more relative feature information from the data than the roulette methods and can generate more valid and accurate measurements. Data from multiple sensors can be combined, which helps remove noise and variability in the signals, providing more precise readings related to blood pressure. In addition, feature selection techniques aid in identifying the most important and informative features, thereby reducing the dimension of the data and increasing the accuracy of prediction models (Dhara et al., 2023). This technique may result in more accurate and reproducible blood pressure measurements (Yao et al., 2022). These are effective in the management of hypertension and also for the management of cardiovascular disease. As technology continues to evolve, deep learning-driven sensor fusion and feature selection techniques could also make blood pressure measurements more accurate and useful in clinical applications (González et al., 2023). In general, DL has the potential to dramatically enhance the quality of healthcare and patient management, particularly in accurately estimating blood pressure rates, as suggested by the insights from this study.

References

[1] Abualkishik, A. Z., Almajed, R., & Almutairi, S. A. (2022). Early detection of cardiovascular diseases using deep learning feature fusion and MRI image analysis. *Fusion: Practice and Applications, 8*, 16-24. https://doi.org/10.54216/FPA.080202

[2] Aliyar Vellameeran, F., & Brindha, T. (2022). A new variant of deep belief network assisted with optimal feature selection for heart disease diagnosis using IoT wearable medical devices. *Computer Methods in Biomechanics and Biomedical Engineering*, *25*(4), 387-411. https://doi.org/10.1080/10255842.2021.1955360

[3] Attivissimo, F., D'Alessandro, V. I., De Palma, L., Lanzolla, A. M. L., & Di Nisio, A. (2023). Non-invasive blood pressure sensing via machine learning. *Sensors (Basel, Switzerland)*, *23*(19), 8342. https://doi.org/10.3390/s23198342

[4] Biswas, D., & Tiwari, A. (2024). Utilizing computer vision and deep learning to detect and monitor insects in real time by analyzing camera trap images. *Natural and Engineering Sciences*, *9*(2), 280-292. https://doi.org/10.28978/nesciences.1575480

[5] Chao, P. C. P., Wu, C. C., Nguyen, D. H., Nguyen, B. S., Huang, P. C., & Le, V. H. (2021). The machine learnings leading the cuffless PPG blood pressure sensors into the next stage. *IEEE Sensors Journal*, *21*(11), 12498-12510. https://doi.org/10.1109/JSEN.2021.3073850

[6] Chauhana, N., & Iqbal, A. (2024). Blockchain-SDN Framework for Secure and Reliable 5G Massive IoT Networks Utilizing Deep Learning for Bandwidth Estimation. *International Academic Journal of Science and Engineering*, *11*(2), 28–32. https://doi.org/10.71086/IAJSE/V11I2/IAJSE1146

[7] Cheng, H., Xiong, J., Chen, Z., & Chen, J. (2023). Deep learning-based non-contact IPPG signal blood pressure measurement research. *Sensors*, *23*(12), 5528. https://doi.org/10.3390/s23125528

[8] Dhara, S. K., Bhanja, N., & Khampariya, P. (2023). Optimal fused feature selection with ensemble learning foratrial fibrillation detection using ECG with enhanced average and subtraction-based optimizer. *Intelligent Decision Technologies*, *17*(4), 1489-1513. https://doi.org/10.3233/IDT-220130

[9] El Attaoui, A., Largo, S., Jilbab, A., & Bourouhou, A. (2021). Wireless medical sensor network for blood pressure monitoring based on machine learning for real-time data classification. *Journal of Ambient Intelligence and Humanized Computing*, *12*(9), 8777-8792. https://doi.org/10.1007/s12652-020-02660-1

[10] Eom, H., Roh, J., Hariyani, Y. S., Baek, S., Lee, S., Kim, S., & Park, C. (2021). Deep learning-based optimal smart shoes sensor selection for energy expenditure and heart rate estimation. *Sensors*, *21*(21), 7058. https://doi.org/10.3390/s21217058

[11] González, S., Hsieh, W. T., & Chen, T. P. C. (2023). A benchmark for machine-learning based non-invasive blood pressure estimation using photoplethysmogram. *Scientific Data*, *10*(1), 149. https://doi.org/10.1038/s41597-023-02020-6

[12] Hassan, M. R., Huda, S., Hassan, M. M., Abawajy, J., Alsanad, A., & Fortino, G. (2022). Early detection of cardiovascular autonomic neuropathy: A multi-class classification model based on feature selection and deep learning feature fusion. *Information Fusion*, *77*, 70-80. https://doi.org/10.1016/j.inffus.2021.07.010

[13] Islam, M. A., Majumder, M. Z. H., Miah, M. S., & Jannaty, S. (2024). Precision healthcare: A deep dive into machine learning algorithms and feature selection strategies for accurate heart disease prediction. *Computers in Biology and Medicine*, *176*, 108432. https://doi.org/10.1016/j.compbiomed.2024.108432

[14] Ismail, S. N. A., Nayan, N. A., Jaafar, R., & May, Z. (2022). Recent advances in non-invasive blood pressure monitoring and prediction using a machine learning approach. *Sensors*, *22*(16), 6195. https://doi.org/10.3390/s22166195

[15] Jijesh, J. J., Shivashankar, & Keshavamurthy. (2021). A supervised learning based decision support system for multi-sensor healthcare data from wireless body sensor networks. *Wireless Personal Communications*, *116*(3), 1795-1813. https://doi.org/10.1007/s11277-020-07762-9

[16] Khan, M. U., Aziz, S., Akram, T., Amjad, F., Iqtidar, K., Nam, Y., & Khan, M. A. (2021). Expert hypertension detection system featuring pulse plethysmograph signals and hybrid feature selection and reduction scheme. *Sensors*, *21*(1), 247. https://doi.org/10.3390/s21010247

[17] Krishnan, R., & Dandekar, G. (2022). Medical Image Analysis for Disease Detection and Categorisation Using Deep Learning Algorithm. *International Academic Journal of Science and Engineering*, *9*(3), 16–20. https://doi.org/10.71086/IAJSE/V9I3/IAJSE0923

[18] Liu, S., Huang, Z., Zhu, J., Liu, B., & Zhou, P. (2024). Continuous blood pressure monitoring using photoplethysmography and electrocardiogram signals by random forest feature selection and GWO-GBRT prediction model. *Biomedical Signal Processing and Control*, *88*, 105354. https://doi.org/10.1016/j.bspc.2023.105354

[19] Mathboob, Y. M., Rahaim, L. A. A., & Ali, A. H. (2024). Healthcare Monitoring-based Internet of Things (IOT). *Journal of Internet Services and Information Security*, *14*(4), 347-359. https://doi.org/10.58346/JISIS.2024.I4.021

[20] Muthu Ganesh, V., & Nithiyanantham, J. (2022). Heuristic-based channel selection with enhanced deep learning for heart disease prediction under WBAN. *Computer Methods in Biomechanics and Biomedical Engineering*, *25*(13), 1429-1448. https://doi.org/10.1080/10255842.2021.2013828

[21] Nancy, A. A., Ravindran, D., Raj Vincent, P. D., Srinivasan, K., & Gutierrez Reina, D. (2022). Iot-cloud-based smart healthcare monitoring system for heart disease prediction via deep learning. *Electronics*, *11*(15), 2292. https://doi.org/10.3390/electronics11152292

[22] Pan, F., He, P., Wang, H., Xu, Y., Pu, X., Zhao, Q., ... & Zheng, D. (2021). Development and validation of a deep learning-based automatic auscultatory blood pressure measurement method. *Biomedical Signal Processing and Control*, *68*, 102742. https://doi.org/10.1016/j.bspc.2021.102742

[23] Qiu, S., Zhao, H., Jiang, N., Wang, Z., Liu, L., An, Y., ... & Fortino, G. (2022). Multi-sensor information fusion based on machine learning for real applications in human activity recognition: State-of-the-art and research challenges. *Information Fusion*, *80*, 241-265. https://doi.org/10.1016/j.inffus.2021.11.006

[24] Rajalakshmi, R., Sivakumar, P., Kumari, L. K., & Selvi, M. C. (2024). A novel deep learning model for diabetes melliuts prediction in IoT-based healthcare environment with effective feature selection mechanism. *The Journal of Supercomputing*, *80*(1), 271-291. https://doi.org/10.1007/s11227-023-05496-6

[25] Saranya, G., & Pravin, A. (2023). A novel feature selection approach with integrated feature sensitivity and feature correlation for improved prediction of heart disease. *Journal of Ambient Intelligence and Humanized Computing*, *14*(9), 12005-12019. https://doi.org/10.1007/s12652-022-03750-y

[26] Schrumpf, F., Frenzel, P., Aust, C., Osterhoff, G., & Fuchs, M. (2021). Assessment of non-invasive blood pressure prediction from ppg and rppg signals using deep learning. *Sensors*, *21*(18), 6022. https://doi.org/10.3390/s21186022

[27] Sikarwar, T. S., Chauhan, A. S., Jain, N., & Mathur, H. (2025). Application of Predictive Analytics in IOT Data Processing. *Indian Journal of Information Sources and Services*, *15*(2), 340–348. https://doi.org/10.51983/ijiss-2025.IJISS.15.2.42.

[28] Veera Boopathy, E., Peer Mohamed Appa, M. A. Y., Pragadeswaran, S., Karthick Raja, D., Gowtham, M., Kishore, R., Vimalraj, P., & Vissnuvardhan, K. (2024). A Data Driven Approach Through IOMT Based Patient Healthcare Monitoring System. *Archives for Technical Sciences*, *2*(31), 9–15. https://doi.org/10.70102/afts.2024.1631.009

[29] Yao, P., Xue, N., Yin, S., You, C., Guo, Y., Shi, Y., ... & Zhao, M. (2022). Multi-dimensional feature combination method for continuous blood pressure measurement based on wrist PPG sensor. *IEEE Journal of Biomedical and Health Informatics*, *26*(8), 3708-3719. https://doi.org/10.1109/JBHI.2022.3167059

[30] Yen, C. T., Liao, J. X., & Huang, Y. K. (2021). Applying a deep learning network in continuous physiological parameter estimation based on photoplethysmography sensor signals. *IEEE sensors journal*, *22*(1), 385-392. https://doi.org/10.1109/JSEN.2021.3126744

[31] Zhao, L., Liang, C., Huang, Y., Zhou, G., Xiao, Y., Ji, N., ... & Zhao, N. (2023). Emerging sensing and modeling technologies for wearable and cuffless blood pressure monitoring. *npj Digital Medicine*, *6*(1), 93. https://doi.org/10.1038/s41746-023-00835-6

Authors Biography

Sakshi Sobti is associated with the Centre of Research Impact and Outcome at Chitkara University, Punjab. Her research focus includes academic performance evaluation, institutional research analytics, and innovation management. She contributes to projects aimed at enhancing research quality, promoting interdisciplinary collaboration, and fostering data-driven strategies within higher education institutions.

Dr. Saurabh Jain is an Associate Professor in the Department of Electronics Engineering at Medi-Caps University, Indore. His research interests include signal processing, embedded systems, and communication technologies. He has published multiple research papers and is actively engaged in teaching, mentoring students, and guiding research projects in electronics and engineering applications.

Dr. Arunkumar Devalapura Thimmappa is an Assistant Professor in the Department of Mechanical Engineering at JAIN (Deemed-to-be University), Bangalore. His research areas include thermal engineering, renewable energy systems, and mechanical design. He is actively involved in mentoring students, leading engineering projects, and publishing research in peer-reviewed journals.

Dr. Shaikh Adil is an Assistant Professor in the Department of Dairy Technology at Parul Institute of Technology, Parul University. His research interests include food technology, dairy processing, and biotechnology applications in food safety. He is involved in academic teaching, research projects, and mentorship, with a focus on improving dairy technology practices and innovation in food sciences.

Lovish Dhingra is affiliated with the Chitkara Centre for Research and Development, Chitkara University, Himachal Pradesh. His research focus includes institutional research strategy, academic analytics, and outcome-based project management. He contributes to initiatives aimed at improving research quality, promoting innovation, and supporting interdisciplinary collaboration in higher education.

Dr.R.S. Ernest Ravindran is an Associate Professor in the Department of Electronics and Communication Engineering at Koneru Lakshmaiah Education Foundation. His research interests include communication systems, signal processing, and embedded electronics. He has published extensively in peer-reviewed journals and is actively involved in mentoring students, leading research projects, and developing innovative solutions in electronics and communication engineering.